



ATHENICA

ENVIRONMENTAL SERVICES INC.

45-09 Greenpoint Avenue
Long Island City, N.Y. 11104

TEM DEPARTMENT

Tel. (718) 784-7490
Fax (718) 784-4085

TEM AHERA COUNTSHEET

CLIENT: WARREN ~~W~~ PANZER	DATE: 10/13/02
PROJECT: NYC DEP/170 WILLIAMS ST.	ANALYST: E.S.
BATCH #: T02-10-079	CLIENT ID #: 03
SAMPLE ID #: T02-10-079-3083	VOLUME (lt): 1350
G.O. AREA (mm ²): 0.013838	GO's TO ANALYZE: 5
MAGNIFICATION: 19K	ACC. VOLTAGE (KeV): 100
INSTRUMENT: JEOL 100CX II	FILTER AREA (mm ²): 385
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

GRI D	GO ID	Str. #	Str. Classf.	L (μm)	W (μm)	Structure sketch	SAE D	EDXA	PHO TO#	ID
B	F7	NSD								
	H6	NSD								
A	C4	NSD								
	G2	NSD								
	D10	NSD								

Analytical Sensitivity (str/cm ²): 0.0041	Eucentricity check:
Concentration (str/cm ²): <0.0041	*Grid acceptance criteria check: (A) ☒ (B) ☒
Concentration (str/mm ²): <14.45	**CHR. # STRUCT.: F M C B
Number of Asbestos Structures: 0.5 ≤ S < 5 μm 0..	**AMPH. # STRUCT.: F M C B
S ≥ 5 μm 0...	Specify AMPH. Type: AM. ACT. TRE.
Number of Non-asbestos Structures: 0.....	CRO. ANTH.

* Replica covers >50%, at least 20 GO's with no overlapping or folded replica, intact GO's have >50% coverage, <50% Undissolved filter, <5% holes or opaque areas, <10% particulate coverage.

** CHR.=chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE.=tremolite, CRO.=crocidolite, ANTH.=anthophyllite.

CITY_21MC100_011080182