

APR-03-2002 14:06

**WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM**

Proficiency Test Report

Page 1 of 2

Lab : 10879 ATC ASSOCIATES INC EPA Lab Code : NY01349
104 EAST 25TH 10TH FLOOR
NEW YORK, NY-10010

Shipment Date : 05-Mar-2001
Closing Date : 19-Apr-2001
Score Date : 15-May-2001
Report Date : 15-May-2001

Shipment : 241 Asbestos In Air, Water, and NonFriable Samples by Electron Microscopy

Analyte	Unit	Sample ID	Result	Method	Mean/Target	Satisfactory Limit	Score
Sample: Asbestos In Air by TEM							
Approval Category: Air and Emissions							
Asbestos In Air by TEM	Asbestos Structures per	6001	316.7	Method Not Specified	366	56 - 716	Satisfactory
28 passed out of 29 reported results.							
Dominant Asbestos Type		6001	Amosite	Method Not Specified	Amosite		Satisfactory
29 passed out of 29 reported results. QUAL							
Asbestos In Air by TEM	Asbestos Structures per	6756	583.3	Method Not Specified	557	155 - 959	Satisfactory
26 passed out of 29 reported results.							
Dominant Asbestos Type		6756	Actinolite	Method Not Specified	Actinolite		Satisfactory
23 passed out of 29 reported results. QUAL							
Sample: Asbestos In Non-Friable Material							
Approval Category: Solid and Hazardous Waste							
Percent Residue	%	6444	15.3	Method Not Specified	14.6	5.21 - 24	Satisfactory
28 passed out of 29 reported results. RESIDUE							

** Indicates NVLAP accredited analyte. Lab Code 200287-0. ELAP has applied to A2LA for PT Provider accreditation.

99%

P.14

CITY_21MC100_008111672

04/03/2002 13:17 FAX 212 350 8306

ATC

006/009

**WADSWORTH CENTER
NEW YORK STATE DEPARTMENT OF HEALTH
ENVIRONMENTAL LABORATORY APPROVAL PROGRAM**

Proficiency Test Report

Page 2 of 2

Lab : 10879 ATC ASSOCIATES INC EPA Lab Code : NY01349
104 EAST 25TH 10TH FLOOR
NEW YORK, NY-10010

Shipment Date : 05-Mar-2001
Closing Date : 19-Apr-2001
Score Date : 15-May-2001
Report Date : 15-May-2001

Shipment : 241 Asbestos in Air, Water, and NonFriable Samples by Electron Microscopy

Analyte	Unit	Sample ID	Result	Method	Mean/Target	Satisfactory Limits	Score
Sample: Asbestos in Non-Friable Material							
Approval Category : Solid and Hazardous Waste							
Percent Asbestos in Residue	%	6444	50.0	Method Not Specified	65.8	9.7 - 122	Satisfactory
		29 passed out of 20 reported results.	ASB_IN_RES				
Percent Asbestos in Residue	%	7935	25.0	Method Not Specified	37.5	4.69 - 95.7	Satisfactory
		27 passed out of 29 reported results.	ASB_IN_RES				
Percent Residue	%	7935	18.5	Method Not Specified	14.8	4.51 - 25.1	Satisfactory
		27 passed out of 29 reported results.	RESIDUE				
Sample: Asbestos in Water by TEM							
Approval Category : Potable Water							
Asbestos in Water by TEM	Millions of Fibers per liter	8910	5.6	Method Not Specified	6.37	D.L. - 14.2	Satisfactory
		23 passed out of 24 reported results.					

** indicates NVLAP accredited analyte. Lab Code 200387-0. ELAP has applied to A2LA for PT Provider accreditation.

APR-03-2002 14:05

99%

P.10

CITY_21MC100_008111673

04/03/2002 10:15 FAX 212 353 9306

ATC

0001/003