

DCT-16-2001 12:05

P.07/11

ambient labs inc. 159 West 25th Street New York, NY 10001 ELAP #10865

REPORT OF AIR SAMPLE ANALYSIS BY TRANSMISSION ELECTRON MICROSCOPY

Project ID : 0211358
 Client Name : J. Resnick & Sons
 Project Name : J. Resnick & Sons Field ID : 01

Sample ID : AEM990 09424 Date Received : 9/20/01
 Sample Description : Lobby of 75 Park Place Security Desk

Filter Composition : MCE Porosity : 0.45 micron Filt Lot # : EX55
 Filter Diameter : 25 mm Effective Filter Area : 385 mm² Air Volume : 1250 ltr
 Total Filter Area Analyzed : 0.0630 mm²

Grid Openings : 7 Avg Grid Opening Area : 0.0090 mm² Grid Lot # : 4083
 Analytical Sensitivity : 0.0049 asbestos struct/cm³

Number of Asbestos Structures :

Chrysotile : 0
 Amphibole : 0

Number of Structure Types :

LENGTH	FIBER	CLUSTER	MATRIX	BUNDLE
0.5 um < L < 5 um	0	0	0	0
L > 5 um	0	0	0	0

A: TOTAL CHRYSOTILE STRUCTURE CONCENTRATION : LT ANYL SENS structure/cm³
 B: TOTAL AMPHIBOLE STRUCTURE CONCENTRATION : LT ANYL SENS structure/cm³
 C: TOTAL ASBESTOS STRUCTURE (A+B) CONCENTRATION : LT ANYL SENS structure/cm³

TOTAL ASBESTOS CONCENTRATION :	IN AIR	LT ANYL SENS	structure/cm ³
	ON FILTER	LT ANYL SENS	structure/mm ²

Micrograph ID # for Reported Structures: None None

Reported for Ambient Labs Inc. by :

Inna Nurik

Analysis Date : 9/20/01

INNA NURIK
 Environmental Analyst

* Overloaded, NA - Analyzed