NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	**Volume (liters):**	1816
Lab Group #:	6800	**Filter Type / Filter Area:**	MCE 385 mm²
Lab Sample #:	10	**Grid Opening Size:**	0 01124
Client Sample #:	13	**Area Examined:**	0 04496 mm²
GRID BOX:	5R6	**Magnification:**	20,000
GRID ID:	5R15	**Accelerating Voltage:**	100 KeV
# Grid Openings:	4	**Analysis Date:**	2/24/02

Analysis Performed By: RP
ROMAN PEYSAKHOV

5R6 / 15

GO No	GO Coord	Structure No	Structure Type F	B	C	M	Length >=05 µ <5µ	>=5µ	ID of Structure √ or N Chrys	Amphibole	Nonasb	Ambig/ND	EDS (Nonadb Elements)
1	I 2	NSD											
2	I 2	〃											
3	I 2	〃											
4	J 2	〃											

Struct = > 0 5µ > 5 1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber
B = Bundle (<3 fibers <1 fiber diameter apart
C = Cluster (>3 Intersections)
M = Matrix
NSD = No Structures Detected

Grid acceptance

<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

NYC-WTC_000190121