TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	**Volume (liters).**	1808
Lab Group #.	6856	**Filter Type / Filter Area:**	MCE 385 mm²
Lab Sample #.	10	**Grid Opening Size:**	0 01124
Client Sample #:	13	**Area Examined:**	0 04496 mm-
GRID BOX:	E-C	**Magnification:**	20 000
GRID ID:	2515	**Accelerating Voltage:**	100 KeV
# Grid Openings:	4	**Analysis Date:**	2/27/02

Analysis Performed By: MARK PEYSAKHOV

GO No	GO Coord	Structure No	Structure Type F	B	C	M	Length >=05 µ <5µ	>=5µ	ID of Structure No of √ Chrys	Amphibole	Nonasb	Ambig/ND	EDS (Nonadb Elements)
25	1 E1	NSD											
	2 D1	- "											
	3 H1	- "											
35	4 M1	- "											

Struct = > 0 5µ > 5 1 Aspect Ratio

√= Positive diffraction pattern properties conforming to aboratory identification criteria observed

√= EDS spectrum conforming to laboratory dentification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart

C = Cluster (>3 Intersections)

M = Matrix

NSD = No Structures Detected

Grid acceptance

<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES

Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					