



TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	Volume (liters):	1824
Lab Group #:	7932	Filter Type / Filter Area:	MCE 385 mm ²
Lab Sample #:	16	Grid Opening Size:	0.01124
Client Sample #:	19	Area Examined:	0.04496 mm ²
GRID BOX:		Magnification:	20,000
GRID ID:		Accelerating Voltage:	100 KeV
# Grid Openings:	4	Analysis Date:	6/1/02

Analysis Performed By:

ROMAN PEYSAKHOV

GO.		Structure No	Structure Type				Length		ID of Structure N or ✓				EDS (Nonadb Elements)
No	Coord		F	B	C	M	>=05 μ <5μ	>=5μ	Chrys	Amphibole	Nonasb	Ambig/ND	
1	Q2	NSD											
2	J2	"											
3	P2	"											
4	I2	"											
5	I2	"											
6													
7													
8													
9													
10													
11													
12													
13													
14													
15													
16													
17													
18													
19													
20													
21													
22													
23													
24													
25													
26													
27													
28													
29													
30													
31													
32													
33													
34													
35													
36													
37													
38													
39													
40													
41													
42													
43													
44													
45													
46													
47													
48													
49													
50													

Struct = > 0.5μ > 5:1 Aspect Ratio

✓ = Positive diffraction pattern properties conforming to laboratory identification criteria observed

✓ = EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart)

C = Cluster (>3 Intersections)

M = Matrix

NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage at least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES

Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5μ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					