

<u>Client Name:</u>	NYC BOE	<u>Volume (liters):</u>	1912
<u>Lab Group #:</u>	6644	<u>Filter Type / Filter Area:</u>	MCE 385 mm ²
<u>Lab Sample #:</u>	7	<u>Grid Opening Size:</u>	0.01124
<u>Client Sample #:</u>	7	<u>Area Examined:</u>	0.04496 mm ²
<u>GRID BOX:</u>	<u>816</u>	<u>Magnification:</u>	20,000
<u>GRID ID:</u>	<u>204307</u>	<u>Accelerating Voltage:</u>	100 KeV
<u># Grid Openings:</u>	4	<u>Analysis Date:</u>	2/12/02


MARK PEYSAKHOV

[illegible]

Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

	F	B	C	M	>5μ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					