

ATTACHMENT 2

ANALYTICAL LABORATORY REPORT

MEM[®] (GRAVIMETRIC) SAMPLES

CITY_21MC100_003056557

Feb 12 02 02:30p

PSC Analytical Services

6109219667

p.8



INDUSTRIAL HYGIENE

ENVIRONMENTAL TESTING

• EPAN/VLAP 101262-0

• NY DOH/ELAC 10903

• NJ DEP 77020

• AIHA ACCREDITATION NO. 100439

• PA DEP 06-353

• CT DPH PH-0238

• NC DENR 599

Analytical Results
for

ATC Associates

WorkOrder: R02020289

Client Reference: 19125-0246

Analyte	Concentration			Limit of Detection (ug)	Test Method	Date Analyzed /Analyst
	(ug)	(mg/m ³)	(ppm)			

Client ID: Dust1 ATC-0206-STY Lab ID: -01A Date Sampled: 2/6/02 Matrix: Filter Air Vol.(L): 1440

Total Particulate Matter 50.0 0.00347 -- 10 N0500 02/11/2002 AJB

Main Lobby

Client ID: Dust2 ATC-0206-STY Lab ID: -02A Date Sampled: 2/6/02 Matrix: Filter Air Vol.(L): 1440

Total Particulate Matter 130 0.00903 -- 10 N0500 02/11/2002 AJB

Bridge entrance

(a) Analysis indicates possible breakthrough; back section result is greater than 0% of the front section result.

General Notes:

<: Less than the indicated limit of detection (LOD).

--: Information not available or not applicable.

Back sections were checked and showed no significant breakthrough.