

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	Volume (liters):	1840
Lab Group #:	6502	Filter Type / Filter Area:	MCE 385 mm ²
Lab Sample #:	10	Grid Opening Size:	0.01124
Client Sample #:	10	Area Examined:	0.04496 mm ²
GRID BOX:	776	Magnification:	20,000
GRID ID:	2938	Accelerating Voltage:	100 KeV
# Grid Openings:	4	Analysis Date:	1/30/02

Analysis Performed By: RA
ROMAN PEYSAKHOV

GO.	Structure No	Structure Type				Length		ID of Structure N or ✓				EDS (Nonadb Elements)	
		F	B	C	M	>=05 μ <5μ	>=5μ	Chrys	Amphibole	Nonasb	Ambig/ND		
2S	1	P1											
	2	I1											
	3	P2											
	4	I2											
3S													

Struct = > 0.5μ > 5:1 Aspect Ratio

✓ = Positive diffraction pattern properties conforming to laboratory identification criteria observed

✓ = EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart)

C = Cluster (>3 Intersections)

M = Matrix

NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate
least 20 unfolded or overlapped grid squares <5%
holes or opaque areas

AMPHIBOLES

Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5μ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					