NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	Volume (liters):	1840
Lab Group #:	6939	Filter Type / Filter Area:	MCE 385 mm²
Lab Sample #:	16	Grid Opening Size:	0.01124
Client Sample #:	16	Area Examined:	0.04496 mm²
GRID BOX:	6PG	Magnification:	20,000
GRID ID:	E05-D	Accelerating Voltage:	100 KeV
# Grid Openings:	4	Analysis Date:	3/5/02

Analysis Performed By: MARK PEYSAKHOV

GO		Structure No	Structure Type				Length		ID of Structure N or √				EDS (Nonadb Elements)
No	Coord		F	B	C	M	>=05 µ <5µ	>=5µ	Chrys	Amphibole	Nonasb	Ambig/ND	
4D	1 P2	4950											
	2 D2	— "											
5-D	3 H1	— "											
	4 H1	— "											

Struct = > 0.5µ > 5:1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart

C = Cluster (>3 Intersections)

M = Matrix

NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded <10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES

Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Amoiguous					
Non-Asbestos					
Non-Identified					

NYC-WTC_000201827