NYC 9/11 Public Portal Document

ATC ASSOCIATES INC. – TEM ANALYSIS

LAB GROUP #: S606 - 10/04/01

Lab Sample No. 60104 | Filter Type MCE 25 mm | Operator MP

Client Spl. No. HSEF-E-10/3-26 | Filter Area 385 mm^2

Instrument I.D. EM 10 CRX | Grid I.D. 432 | Disk No.

Screen Magnification 16,400X | (GO) Area 0.01124 mm^2 | Comments

Acc. Voltage 80 kV | No. GO to be Analyzed 4 | (3.147 Struct. Corspd. to 70 s/mm^2.)

GO. No.	GO. Coord.	Structure No.	Structure Type F	B	C	M	Length ≥0.5 um < 5 um	Length ≥ 5um	I.D. of Structure Chrys.	Amphibole	Nonasb.	N or ✓ Ambig./ND	EDS (Nonasb. Elements)
1	E6	NSD											
2	R6	— " —											
3	X7	— " —											
4	Q7	— " —											

3P

3R

Photo No. (from 1 in 5 asbestos-containing samples)

Identify all structures using EDS

Check diffraction for asbestos structures up to 70s/mm (1st 4)

Store all unrepeated spectra

Struct. = ≥ 0.5um & > 5:1 Aspect Ratio

N= Not Observed (EDS spectrum or diffraction pattern)

✓= Positive diffraction pattern properties conforming to laboratory identification criteria observed.

For chrysotile 0.73nm spacing for (002) reflections. 0.53nm spacing for the layer line repeat, streaking of (110) and (130) reflections.


For amphiboles: 0.53nm layer line repeat, and approx. 3.5 ratio of non-twinned layer line spot spacing to layer line repeat distance (0.53nm).

✓= EDS spectrum conforming to laboratory identification criteria observed.

F= Single Fiber (1-3 intersections = 1-3 fibers)
B= Bundle (>3 fibers <1 fiber diameter apart)
C= Cluster (>3 intersections)
M= Matrix
NSD= No Structures Detected

L\DATA\PRIVATE\MIKEC\CADD\TEM2 9/22/96

Grid acceptance:
<50% undissolved filter, overlapped or folded
<10% particulate coverage >80% coverage
at least 20 unfolded or overlapped grid squares
<2% holes or opaque areas



Stopping Rules:
>50 s/G.O. or in 1st 4 G.O.

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	≥ 5um
Chrysotile					
Amphibole					
Ambiguous					
Non-asbestos					
Non-Identified					

CITY_21MC100_003274962

NYC-WTC_000202658