

09/30/2001 05:00

212-290-0050

EMSL MANHATTAN

PAGE 01/02

New York Lab

307 West 38th Street, New York, NY 10018

Phone: 212-290-0051 Fax: 212-250-0050 Email:

EMSL

Attn: JLC Environmental Consultants, Inc
200 Park Ave South
Suite 1108
New York, NY 10003

Fax: (212) 420-8092

Phone: 212-420-1119

Project: WTC/DEP/9/28/01

Customer ID: JLCC50

Customer PO:

Received: 09/29/01 9:50 PM

EMSL Order: 030105133

EMSL Project ID:


Analysis Date: 9/29/01

**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM) Performed by
EPA 40 CFR Part 763 Final Rule (AHERA)**

Sample	Volume	Area Analyzed	Non Asb	Asbestos Type(s)	# Structure 0.5µ - 5µ	Analytical Sensitivity (S/cc)	Asbestos Concentration (S/mm ²) (S/cc)	Notes
302/005 030105133-0001	525.00	0.1250	0	None Detected		0.0059	<8.00 <0.0059	Volume is below recommended limit for method.
302/006 030105133-0002	525.00	0.1250	0	Chrysotile	5 5	0.0059	80 0.0567	Volume is below recommended limit for method.
302/007 030105133-0003	525.00			Overloaded				Volume is below recommended limit for method.

Ghayath Elias

Analyst



Laboratory Manager
or other approved signatory

Disclaimer: The laboratory is not responsible for data reported in structures, which is dependent on volume collected by non-laboratory personnel. This lab is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the sample reported above. Quality control data (including 95% confidence limits and laboratory and analyst accuracy and precision) is available upon request.

Accredited for NVLAP PLM/TEM #101048-9, NY State ELAP #11506

TEM/AHERA-1

1

CITY_21MC100_003074447