

New York Lab

307 West 38th Street, New York, NY 10018

Phone: 212-290-0051 Fax: 212-290-0058 Email:

EMSL

Attn: NYC Department of Environmental Protection
96-05 Horace Harding Expressway
Queens, NY 11368

Fax: (718) 595-6307 Phone: 718-595-6291

Project: WTC

Customer ID: +03nycdep
Customer PO:
Received: 10/15/01 6:22 PM
EMSL Order: 030106254
EMSL Project ID:


Analysis Date: 10/15/2001

**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM) Performed by
EPA 40 CFR Part 763 Final Rule (AHERA)**

Sample	Volume	Area Analyzed	Non Asb	Asbestos Type(s)	# Structure 0.5µ - 5µ	Analytical Sensitivity (S/cc)	Asbestos Concentration (S/mm ³) (S/cc)	Notes
802/009 030106254-0001	525.00	0.1500	2	None Detected		0.0049	<6.67 <0.0049	Volume is below recommended limit for method.

Jose Arriaga

Analyst



Laboratory Manager
or other approved signatory

Disclaimer: The laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This lab is only responsible for data reported in structures/mm³. This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analyst accuracy and precision) is available upon request.

Accredited for NVLAP PLM/TEM #101048-9, NY State ELAP #11506

TEMAHERA-1

1