

SEVERN TRENT LABORATORIES, INC.

PRELIMINARY DATA SUMMARY

The results shown below may still require additional laboratory review and are subject to change. Actions taken based on these results are the responsibility of the data user.

Lot #: G2E240264
STL Edison
New York
Project Number: W680
Date Reported: 5/31/02
PAGE 1

PARAMETER	RESULT	REPORTING LIMIT	UNITS	ANALYTICAL METHOD
-----------	--------	--------------------	-------	----------------------

Client Sample ID: W680-351808

Sample #: 001 Date Sampled: 05/22/02 13:00 Date Received: 05/24/02 Matrix: WATER

SUMP WATER

Dioxins/Furans, HRGC/HRMS

In Review

2,3,7,8-TCDD	ND	2.5	pg/L	EPA-5 1613B
1,2,3,7,8-PeCDD	ND	5.7	pg/L	EPA-5 1613B
1,2,3,4,7,8-HxCDD	ND	3.5	pg/L	EPA-5 1613B
1,2,3,6,7,8-HxCDD	ND	3.7	pg/L	EPA-5 1613B
1,2,3,7,8,9-HxCDD	ND	3.3	pg/L	EPA-5 1613B
1,2,3,4,6,7,8-HpCDD	ND	9.7	pg/L	EPA-5 1613B
OCDD	ND	41	pg/L	EPA-5 1613B
2,3,7,8-TCDF	ND	2.9	pg/L	EPA-5 1613B
1,2,3,7,8-PeCDF	ND	3.6	pg/L	EPA-5 1613B
2,3,4,7,8-PeCDF	ND	4.1	pg/L	EPA-5 1613B
1,2,3,4,7,8-HxCDF	ND	1.8	pg/L	EPA-5 1613B
1,2,3,6,7,8-HxCDF	ND	1.7	pg/L	EPA-5 1613B
2,3,4,6,7,8-HxCDF	ND	1.5	pg/L	EPA-5 1613B
1,2,3,7,8,9-HxCDF	ND	1.4	pg/L	EPA-5 1613B
1,2,3,4,6,7,8-HpCDF	ND	3.6	pg/L	EPA-5 1613B
1,2,3,4,7,8,9-HpCDF	ND	2.5	pg/L	EPA-5 1613B
OCDF	ND	4.7	pg/L	EPA-5 1613B

Client Sample ID: W680-351809

Sample #: 002 Date Sampled: 05/22/02 12:30 Date Received: 05/24/02 Matrix: WATER

DEWATERING WELL

Dioxins/Furans, HRGC/HRMS

In Review

2,3,7,8-TCDD	9.7 J		pg/L	EPA-5 1613B
1,2,3,7,8-PeCDD	ND	6.4	pg/L	EPA-5 1613B
1,2,3,4,7,8-HxCDD	ND	4.6	pg/L	EPA-5 1613B
1,2,3,6,7,8-HxCDD	ND	4.9	pg/L	EPA-5 1613B
1,2,3,7,8,9-HxCDD	ND	4.4	pg/L	EPA-5 1613B
1,2,3,4,6,7,8-HpCDD	ND	3.0	pg/L	EPA-5 1613B
OCDD	ND	7.0	pg/L	EPA-5 1613B
2,3,7,8-TCDF	ND	4.3	pg/L	EPA-5 1613B
1,2,3,7,8-PeCDF	ND	3.7	pg/L	EPA-5 1613B
2,3,4,7,8-PeCDF	ND	5.5	pg/L	EPA-5 1613B
1,2,3,4,7,8-HxCDF	ND	2.4	pg/L	EPA-5 1613B
1,2,3,6,7,8-HxCDF	ND	2.2	pg/L	EPA-5 1613B
2,3,4,6,7,8-HxCDF	ND	1.9	pg/L	EPA-5 1613B
1,2,3,7,8,9-HxCDF	ND	1.8	pg/L	EPA-5 1613B

(Continued on next page)

SEVERN TRENT LABORATORIES, INC.
PRELIMINARY DATA SUMMARY

The results shown below may still require additional laboratory review and are subject to change. Actions taken based on these results are the responsibility of the data user.

Lot #: G2E240264 **STL Edison** **PAGE 2**
 New York **Date Reported: 5/31/02**
 Project Number: W680

<u>PARAMETER</u>	<u>RESULT</u>	<u>REPORTING</u> <u>LIMIT</u>	<u>UNITS</u>	<u>ANALYTICAL</u> <u>METHOD</u>
------------------	---------------	----------------------------------	--------------	------------------------------------

Client Sample ID: W680-351809

Sample #: 002 Date Sampled: 05/22/02 12:30 Date Received: 05/24/02 Matrix: WATER

Dioxins/Furans, HRGC/HRMS

1,2,3,4,6,7,8-HpCDF	ND	3.1	pg/L	EPA-5 1613B	In Review
1,2,3,4,7,8,9-HpCDF	ND	3.6	pg/L	EPA-5 1613B	
OCDF	ND	4.8	pg/L	EPA-5 1613B	

J Estimated result. Result is less than the reporting limit.