



ATHENICA

ENVIRONMENTAL SERVICES INC.

45-09 Greenpoint Avenue
Long Island City, N.Y. 11104

TEM DEPARTMENT

Tel. (718) 784-7490
Fax (718) 784-4085

TEM AHERA COUNTSHEET

CLIENT : WARREN & PANZER	DATE : 12/2/02
PROJECT : NYC DEP / 300 ALBANY ST.	ANALYST : EMS ES
BATCH # : T02-12-004	CLIENT ID # : 03
SAMPLE ID # : T02-12-004-4346	VOLUME (lt) : 1800
G.O. AREA (mm ²) : 0.013433	GO's TO ANALYZE : 4
MAGNIFICATION : 19K	ACC. VOLTAGE (KeV) : 100
INSTRUMENT : JEOL 100CX II	FILTER AREA (mm ²) : 385
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

[illegible]

Analytical Sensitivity (str/cm ³):	0.0048	Eucentricity check:	
Concentration (str/cm ³):	<0.0048	*Grid acceptance criteria check:	(A)... (B)...
Concentration (str/mm ²):	<18.61	**CHR. # STRUCT.:	F M C B
Number of Asbestos Structures:	0.5 ≤ S < 5 μm 0...	**AMPH. # STRUCT.:	F M C B
	S ≥ 5 μm 0...	Specify AMPH. Type:	AM. ACT. TRE.
Number of Non-asbestos Structures:	0...	CRO. ANTH.	

*Replica covers >50% at least 20 GO's with no overlapping or folded replica, intact GO's have >50% coverage, <50% Undissolved filter, <5% holes or opaque areas, <10% particulate coverage

** CHR = chrysotile, AMPH = amphibole, AM = amesite, ACT = actinolite, TRF = tremolite, CRO = crocidolite, ANTH = anthophyllite