TEM Asbestos (air) Count Sheet

Client Name:	Warren & Panzer	**Volume (liters):**	1440
Lab Group #:	8313	**Filter Type / Filter Area:**	MCE 385 mm²
Lab Sample #:	2	**Grid Opening Size:**	0.01124
Client Sample #:	02	**Area Examined:**	0.0562 mm²
GRID BOX:	1276	**Magnification:**	20,000
GRID ID:	EF10R	**Accelerating Voltage:**	100 KeV
# Grid Openings:	5	**Analysis Date:**	6/27/02

Analysis Performed By: MARK PEYSAKHOV

GO. No	GO. Coord	Structure No	Structure Type F	B	C	M	Length >=05 µ <5µ	>=5µ	ID of Structure N or √ Chrys	Amphibole	Nonasb	Ambig/ND	EDS (Nonadb Elements)
1	01	1			↵			↵	↵				
		2	↵				↵		↵				
		3	↵					↵	↵				
		4	↵				↵		↵				
		5				↵	↵		↵				
2	01	NSD											
3	01	— ↵											
4	XQ	— ↵											
10R	5 QE	— ↵											

Struct = > 0.5µ .> 5:1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber
B = Bundle (<3 fibers <1 fiber diameter apart
C = Cluster (>3 Intersections)
M = Matrix
NSD = No Structures Detected

Grid acceptance:
<50% undissolved filter, overlapped or folded <10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					