



ATHENICA

ENVIRONMENTAL SERVICES INC.



45-09 Greenpoint Avenue
Long Island City, N.Y. 11104

TEM DEPARTMENT

Tel. (718) 784-7490
Fax (718) 784-4085

TEM AHERA COUNTSHEET

CLIENT : <u>QA/QC</u>	DATE : <u>8/31/06</u>
PROJECT :	ANALYST :
BATCH # : <u>702.08-156</u>	CLIENT ID # :
SAMPLE ID # : <u>702.08-156-2022 Sept</u>	VOLUME (lt) : <u>1560</u>
G.O. AREA (mm ²) : <u>0.012422</u>	GO's TO ANALYZE : <u>4</u>
MAGNIFICATION : <u>196</u>	ACC. VOLTAGE (KeV) : <u>120</u>
INSTRUMENT : <u>JEOL 100CX II</u>	FILTER AREA (mm ²) : <u>385</u>
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

GRI D	GO ID	Str. #	Str. Classf.	L (μm)	W (μm)	Structure sketch	SAE D	EDXA	PHO TO#	ID
A	B4	N>D								
	E5	N>D								
B	H3	1	A F	8.0	0.5		chw	Mgs. 12		chw
			B F	0.75	0.05		chw	-11-		chw
	DC	1	B	8.0	0.20		chw	-11-		chw

Analytical Sensitivity (str/cm ³): <u>0.0050</u>	Eucentricity check : <u>✓</u>
Concentration (str/cm ³): <u>0.0149</u>	*Grid acceptance criteria check : (A)...✓ (B)...
Concentration (str/mm ²): <u>60.38</u>	**CHR. # STRUCT.: F M C B
Number of Asbestos Structures : 0.5 ≤ S < 5 μm : <u>0.1</u>	**AMPH. # STRUCT.: F M C B
S ≥ 5 μm : <u>0.2</u>	Specify AMPH. Type : AM. ACT. TRE.
Number of Non-asbestos Structures: <u>0</u>	CRO. ANTH.

*Replica covers >50%, at least 20 GO's with no overlapping or folded replica, intact GO's have >50% coverage, <50% Undissolved filter, <5% holes or opaque areas, <10% particulate coverage.

** CHR.=chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE=tremolite, CRO.=crocidolite, ANTH.=anthophyllite.

ENVIRONMENTAL SERVICES INC.

Tel. (718) 784-7480
Fax (718) 784-4085

CLIENT: WARREN #PANZER	DATE: 8/31/02
PROJECT: NYC DEP/20 Broad St.	ANALYST: E.S.
BATCH #: T02-08-156	CLIENT ID #: 04
SAMPLE ID #: T02-08-156-2025	VOLUME (lt): 1560
G.O. AREA (mm ²): 0.012422	GO's TO ANALYZE: 4
MAGNIFICATION: (9K)	ACC. VOLTAGE (KeV): 100
INSTRUMENT: JEOL 100CX II	FILTER AREA (mm ²): 385
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

[illegible]

Analytical Sensitivity (str/cm ³):	0.0050	Eucentricity check:	
Concentration (str/cm ³):	<0.0050	*Grid acceptance criteria check:	(A) ☒ (B) ☒
Concentration (str/mm ²):	<20.13	**CHR. # STRUCT.:	F M C B
Number of Asbestos Structures:	0.5 ≤ S < 5 μm 0...	**AMPH. # STRUCT.:	F M C B
	S ≥ 5 μm 0...	Specify AMPH. Type:	AM. ACT. TRE.
Number of Non-asbestos Structures:	0...	CRO. ANTH.	

** CHR.=chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE.=tremolite, CRO.=crocidolite, ANTH.=anthophyllite.

ATHENICA
ENVIRONMENTAL SERVICES INC.

45-08 Greenpoint Avenue
Long Island City, N.Y. 11104

TEM DEPARTMENT

Tel. (718) 784-7490
Fax (718) 784-4085

TEM AHERA COUNTSHEET

CLIENT: <u>WARREN FRANZER</u>	DATE: <u>8/31/02</u>
PROJECT: <u>NYC DEP / 20 BROAD ST.</u>	ANALYST: <u>E.S.</u>
BATCH #: <u>T02-04-156</u>	CLIENT ID #: <u>05</u>
SAMPLE ID #: <u>T02-04-156-2026</u>	VOLUME (li): <u>1560</u>
G.O. AREA (mm ²): <u>0.012422</u>	GO's TO ANALYZE: <u>4</u>
MAGNIFICATION: <u>19K</u>	ACC. VOLTAGE (KeV): <u>100</u>
INSTRUMENT: <u>JEOL 100CX II</u>	FILTER AREA (mm ²): <u>385</u>
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

GRI D	GO ID	Str. #	Str. Classf.	L (μm)	W (μm)	Structure sketch	SAE D	EDXA	PHO TO#	ID
A	E5	MSD								
	G9	MSD								
B	E7	MSD								
	H3	MSD								

Analytical Sensitivity (str/cm ²): <u>0.0050</u>	Eucentricity check:
Concentration (str/cm ²): <u><0.0050</u>	*Grid acceptance criteria check: (A) ☒ (B) ☐
Concentration (str/mm ²): <u><20.13</u>	**CHR. # STRUCT.: F M C B
Number of Asbestos Structures: 0.5 ≤ S < 5 μm ... <u>0</u>	**AMPH. # STRUCT.: F M C B
S ≥ 5 μm <u>0</u>	Specify AMPH. Type: AM. ACT. TRE.
Number of Non-asbestos Structures: <u>0</u>	CRO. ANTH.

* Replica covers >50%, at least 20 GO's with no overlapping or folded replica, intact GO's have >50% coverage, <50% Undissolved filter, <5% holes or opaque areas, <10% particulate coverage.
 ** CHR.=chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE.=tremolite, CRO.=crocidolite, ANTH.=anthophyllite