

ENVIRONMENTAL SERVICES INC.

Tel. (718) 784-7480
Fax (718) 784-4085

CLIENT : WARREN & PANZER	DATE : 11/12/02
PROJECT : NCDOP/100 CHURCH ST.	ANALYST : E.S.
BATCH # : T02-11-061	CLIENT ID # : 01
SAMPLE ID # : T02-11-061-4010	VOLUME (lt) : 1500
G.O. AREA (mm ²) : 0.013433	GO's TO ANALYZE : 4
MAGNIFICATION : 19K	ACC. VOLTAGE (KeV) : 100
INSTRUMENT : JEOL 100CX II	FILTER AREA (mm ²) : 385
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

[illegible]

Analytical Sensitivity (str/cm ³):	0.0048	Eucentricity check:	
Concentration (str/cm ³):	<0.0048	*Grid acceptance criteria check:	(A)....(B)....
Concentration (str/mm ²):	<18.61	**CHR. # STRUCT.:	F M C B
Number of Asbestos Structures:	0.5 ≤ S < 5 μm 0...	**AMPH. # STRUCT.:	F M C B
	S ≥ 5 μm 0....	Specify AMPH. Type:	AM. ACT. TRE.
Number of Non-asbestos Structures:	 0....	CRO. ANTH.	

** CHR.=chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE.=tremolite, CRO.=crocidolite, ANTH.=anthophyllite