

Tel. (718) 784-7490
Fax (718) 784-4085

TEM AHERA COUNTSHEET

CLIENT: Warren & Panzer	DATE: 8/30/02
PROJECT: NYCOEP/75 Maiden Lane	ANALYST: E.S.
BATCH #: 702-08-155	CLIENT ID #: 03
SAMPLE ID #: 702-08-155-2019	VOLUME (lt): 1530
G.O. AREA (mm ²): 0.012422	GO's TO ANALYZE: 5
MAGNIFICATION: 194	ACC. VOLTAGE (KeV): 100
INSTRUMENT: JEOL 100CX II	FILTER AREA (mm ²): 385
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

[illegible]

Analytical Sensitivity (str/cm ²):	Eucentricity check :
Concentration (str/cm ²):	*Grid acceptance criteria check : (A)..... (B).....
Concentration (str/mm ²):	**CHR. # STRUCT.: F M C B
Number of Asbestos Structures : 0.5 ≤ S < 5 μm	**AMPH. # STRUCT.: F M C B
S ≥ 5 μm	Specify AMPH. Type : AM. ACT. TRE.
Number of Non-asbestos Structures:	CRO. ANTH.

* Replica covers >50%, at least 20 GO's with no overlapping or folded replica, intact GO's have >50% coverage, <50% Undissolved filter, <5% holes or opaque areas, <10% particulate coverage.

** CHR.=chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE.=tremolite, CRO.=crocidolite, ANTH.=anthophyllite.