

Tel. (718) 784-7490
Fax (718) 784-4085

TEM AHERA COUNTSHEET

CLIENT: NAPPEN & PANZER	DATE: 9/8/02
PROJECT: NYCDER/75 MAIDEN LN.	ANALYST: E.S.
BATCH #: T02-09-021	CLIENT ID #: 02
SAMPLE ID #: T02-09-021-2170	VOLUME (lt): 1440
G.O. AREA (mm ²): 0.013771	GO's TO ANALYZE: 4
MAGNIFICATION: 19K	ACC. VOLTAGE (KeV): 100
INSTRUMENT: JEOL 100CX II	FILTER AREA (mm ²): 385
ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210, OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E	

[illegible]

Analytical Sensitivity (str/cm ³):	0.0049	Eucentricity check:	✓
Concentration (str/cm ³):	0.0049	*Grid acceptance criteria check:	(A) ✓ (B) ✓
Concentration (str/mm ²):	19.15	**CHR. # STRUCT.:	F 0 M 0 C 0 B 0
Number of Asbestos Structures:	0.5 ≤ S < 5 μm 0.1	**AMPH. # STRUCT.:	F M C B
	S ≥ 5 μm 0.	Specify AMPH. Type:	AM. ACT. TRE.
Number of Non-asbestos Structures:	0.	CRO. ANTH.	

* Replica covers >50%, at least 20 GO's with no overlapping or folded replica, intact GO's have >50% coverage, <50% Undissolved filter, <5% holes or opaque areas, <10% particulate coverage.

** CHR.=chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE.=tremolite, CRO.=crocidolite, ANTH.=anthophyllite.