NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	Warren & Panzer	**Volume (liters):**	1310
Lab Group #:	8315	**Filter Type / Filter Area:**	MCE 385 mm²
Lab Sample #:	1	**Grid Opening Size:**	0.01124
Client Sample #:	001	**Area Examined:**	0.06744 mm²
GRID BOX:	1277	**Magnification:**	20,000
GRID ID:	3E6	**Accelerating Voltage:**	100 KeV
# Grid Openings:	6	**Analysis Date:**	6/28/02

Analysis Performed By: MARK PEYSAKHOV

GO.		Structure	Structure Type				Length		ID of Structure N or √				EDS
	No Coord	No	F	B	C	M	>=05 µ <5µ	>=5µ	Chrys	Amphibole	Nonasb	Ambig/ND	(Nonadb Elements)
3E	1 D1	NSD											
	2 C1	—											
	3 B5	—											
	4 C1	—											
36	5 B1	—											
	6 D1	—											

Struct – > 0.5µ > 5:1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber
B = Bundle (<3 fibers <1 fiber diameter apart
C = Cluster (>3 Intersections)
M = Matrix
NSD = No Structures Detectec

Grid acceptance:

<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

CITY_DEP_63WALL_MCGIVNEY_SUBP_000004485

NYC-WTC_000206571