NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	Cole	**Volume (liters):**	630
Lab Group #:	7692	**Filter Type / Filter Area:**	MCE 385 mm²
Lab Sample #:	3	**Grid Opening Size:**	0.01124
Client Sample #:	3	**Area Examined:**	0.1124 mm²
GRID BOX:	1122	**Magnification:**	20,000
GRID ID:	~~EE~~ 12B	**Accelerating Voltage:**	100 KeV
# Grid Openings:	10	**Analysis Date:**	5/10/02

Analysis Performed By: RP
ROMAN PEYSAKHOV

GO.		Structure	Structure Type				Length		ID of Structure N or √				EDS
No	Coord	No	F	B	C	M	>=05 µ <5µ	>=5µ	Chrys	Amphibole	Nonasb	Ambig/ND	(Nonadb Elements)
1	V2	NSD									ND		
2	P2	9									9		
3	I2	9									11		
4	J2	11									11		
5	Q2	11									11		
6	Q2	11									11		
7	I2	1				√	√		√		11		
8	P2	2	√				√		√		11		
		3			√			√	√		11		
9	Q3	4	√				√		√		11		
10	J2	NSD											

Struct = > 0.5µ > 5:1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber
B = Bundle (<3 fibers <1 fiber diameter apart
C = Cluster (>3 Intersections)
M = Matrix
NSD = No Structures Detected

Grid acceptance:
<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

NYC-WTC_000208563