NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	Cole	**Volume (liters):**	450
Lab Group #:	7815	**Filter Type / Filter Area:**	MCE 385 mm²
Lab Sample #:	2	**Grid Opening Size:**	0.01124
Client Sample #:	2	**Area Examined:**	0.1124 mm²
GRID BOX:	1157	**Magnification:**	20,000
GRID ID:	7HC	**Accelerating Voltage:**	100 KeV
# Grid Openings:	10	**Analysis Date:**	5/21/02

Analysis Performed By: MARK PEYSAKHOV

GO. No	GO. Coord	Structure No	Structure Type F	B	C	M	Length >=05 µ <5µ	Length >=5µ	ID of Structure N or √ Chrys	Amphibole	Nonasb	Ambig/ND	EDS (Nonadb Elements)
1	A	#20											
2	D1	— c											
3	A1	— c											
4	E1	— c											
5	D1	— c											
6	B2	— c											
7	E2	— c											
8	J2	— c											
9	B2	— c											
10	X2	— c											

Struct = > 0.5µ > 5:1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber
B = Bundle (<3 fibers <1 fiber diameter apart
C = Cluster (>3 Intersections)
M = Matrix
NSD = No Structures Detected

Grid acceptance:
<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

CITY_21MC100_010194525

NYC-WTC_000208579