NYC 9/11 Public Portal Document

EMSL Analytical, Inc.

208 Stonehinge Lane
Carle Place, NY 11514
Phone: (516) 997-7251 Fax: (516) 997-7528

Attn.:
Commercial Lab Enterprise
270 Swinton Avenue
Bronx, NY 10465

Friday, November 02, 2001

Ref Number: LI0116584

Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES ≥ 0.5µ - < 5µ	# STRUCTURES ≥ 5µ	# STRUCTURES Non-Asbestos	Area Analyzed (mm²)	Analytical Sensitivity (S/cc)	Asbestos Concentration (S/mm²)	Asbestos Concentration (S/cc)
469	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
470	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
471	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
472	1390.00	None Detected			0	0.0635	0.0044	<15.75	<0.0044
473	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
474	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045

Earl Tracey
Analyst

Approved Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data including 95% confidence limits and laboratory and analysts' accuracy and precision is available upon request.

EMSL is NVLAP Accredited #101048-0, NJ ELAP #11469, CA ELAP #2360, CT #PH-0200

NYC-WTC_000209611

NYC 9/11 Public Portal Document

EMSL Analytical, Inc.

208 Stonehinge Lane
Carle Place, NY 11514
Phone: (516) 997-7251 *Fax: (516) 997-7528*

Attn.:
Commercial Lab Enterprise
270 Swinton Avenue
Bronx, NY 10465

Friday, November 02, 2001

Ref Number: LI0116584

Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES ≥ 0.5µ - < 5µ	# STRUCTURES ≥ 5µ	# STRUCTURES Non-Asbestos	Area Analyzed (mm²)	Analytical Sensitivity (S/cc)	Asbestos Concentration (S/mm²)	Asbestos Concentration (S/cc)
475	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
476	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
477	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
478	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
479	1340.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
480	1340.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045

Earl Tracey
Analyst

Approved Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

CITY_21MC100_002859262

NYC-WTC_000209612

NYC 9/11 Public Portal Document

EMSL Analytical, Inc.

208 Stonehinge Lane
Carle Place, NY 11514
Phone: (516) 997-7251 Fax: (516) 997-7528

Attn.:
Commercial Lab Enterprise
270 Swinton Avenue
Bronx, NY 10465

Friday, November 02, 2001

Ref Number: LI0116584

Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES ≥ 0.5µ - < 5µ	# STRUCTURES ≥ 5µ	# STRUCTURES Non-Asbestos	Area Analyzed (mm²)	Analytical Sensitivity (S/cc)	Asbestos Concentration (S/mm²)	Asbestos Concentration (S/cc)
481	1340.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
482	1370.00	None Detected			0	0.0635	0.0044	<15.75	<0.0044
483	1200.00	None Detected			0	0.0762	0.0042	<13.12	<0.0042
484	1800.00	None Detected			0	0.0508	0.0042	<19.69	<0.0042
485	1800.00	None Detected			0	0.0508	0.0042	<19.69	<0.0042
486	1800.00	None Detected			0	0.0508	0.0042	<19.69	<0.0042

Earl Tracey
Analyst

Approved Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

Accredited by NVLAP, Lab Code [illegible], NJ LAP 11565, CA LAP [illegible]

CITY_21MC100_002859263

NYC-WTC_000209613

NYC 9/11 Public Portal Document

EMSL Analytical, Inc.

208 Stonehinge Lane
Carle Place, NY 11514
Phone: (516) 997-7251 Fax: (516) 997-7528

Attn.:
Commercial Lab Enterprise
270 Swinton Avenue
Bronx, NY 10465

Friday, November 02, 2001

Ref Number: LI0116584

Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES ≥ 0.5µ - < 5µ	# STRUCTURES ≥ 5µ	# STRUCTURES Non-Asbestos	Area Analyzed (mm²)	Analytical Sensitivity (S/cc)	Asbestos Concentration (S/mm²)	Asbestos Concentration (S/cc)
487	1800.00	None Detected			0	0.0508	0.0042	<19.69	<0.0042
488	1790.00	None Detected			0	0.0508	0.0042	<19.69	<0.0042
489	1750.00	None Detected			0	0.0508	0.0043	<19.69	<0.0043
490	1750.00	None Detected			0	0.0508	0.0043	<19.69	<0.0043
491	1750.00	None Detected			0	0.0508	0.0043	<19.69	<0.0043
492	1750.00	None Detected			0	0.0508	0.0043	<19.69	<0.0043

Earl Tracey
Analyst

Approved
Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

Lab for NVLAP, LAP, LLC, Lab Code 101048-0, NY PLAP, 11459, CA ELAP, 2121, CT PH-0565

NYC-WTC_000209614

NYC 9/11 Public Portal Document

EMSL Analytical, Inc.

208 Stonehinge Lane
Carle Place, NY 11514
Phone: (516) 997-7251 Fax: (516) 997-7528

Attn.:
Commercial Lab Enterprise
270 Swinton Avenue
Bronx, NY 10465

Friday, November 02, 2001

Ref Number: LI0116584

Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES ≥ 0.5µ - < 5µ	# STRUCTURES ≥ 5µ	# STRUCTURES Non-Asbestos	Area Analyzed (mm²)	Analytical Sensitivity (S/cc)	Asbestos Concentration (S/mm²)	Asbestos Concentration (S/cc)
493	1800.00	None Detected			0	0.0508	0.0042	<19.69	<0.0042
494	1750.00	None Detected			0	0.0508	0.0043	<19.69	<0.0043
495	1750.00	None Detected			0	0.0508	0.0043	<19.69	<0.0043
496	1750.00	None Detected			0	0.0508	0.0043	<19.69	<0.0043
497	1750.00	None Detected			0	0.0508	0.0043	<19.69	<0.0043
498	1750.00	None Detected			0	0.0508	0.0043	<19.69	<0.0043

Earl Tracey
Analyst

Approved Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

NVLAP LAP, LLC #101048-0, NY ELAP 11459, CA ELAP 2535, CT PH-0264

CITY_21MC100_002859265

NYC-WTC_000209615

NYC 9/11 Public Portal Document

EMSL Analytical, Inc.

208 Stonehinge Lane
Carle Place, NY 11514
Phone: (516) 997-7251 Fax: (516) 997-7528

Attn.:
Commercial Lab Enterprise
270 Swinton Avenue
Bronx, NY 10465

Friday, November 02, 2001

Ref Number: LI0116584

Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES ≥ 0.5µ - < 5µ	# STRUCTURES ≥ 5µ	# STRUCTURES Non-Asbestos	Area Analyzed (mm²)	Analytical Sensitivity (S/cc)	Asbestos Concentration (S/mm²)	Asbestos Concentration (S/cc)
499	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
500	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
501	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
502	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
503	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
504	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045

Earl Tracey
Analyst

Approved Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

Accredited for NVLAP PLM Test #101048-0, NY ELAP 11486, CA ELAP 2389, CT PH-0284

CITY_21MC100_002859266

NYC-WTC_000209616

NYC 9/11 Public Portal Document

EMSL Analytical, Inc.

208 Stonehinge Lane
Carle Place, NY 11514
Phone: (516) 997-7251 Fax: (516) 997-7528

Attn.:
Commercial Lab Enterprise
270 Swinton Avenue
Bronx, NY 10465

Friday, November 02, 2001

Ref Number: LI0116584

Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES ≥ 0.5µ - < 5µ	# STRUCTURES ≥ 5µ	# STRUCTURES Non-Asbestos	Area Analyzed (mm²)	Analytical Sensitivity (S/cc)	Asbestos Concentration (S/mm²)	Asbestos Concentration (S/cc)
505	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
506	1350.00	None Detected-			0	0.0635	0.0045	<15.75	<0.0045
507	1360.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
508	1360.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
509	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
510	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045

Earl Tracey
Analyst

Approved Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analytical accuracy and precision) is available upon request.

NYC-WTC_000209617

NYC 9/11 Public Portal Document

EMSL Analytical, Inc.

208 Stonehinge Lane
Carle Place, NY 11514
Phone: (516) 997-7251 *Fax: (516) 997-7528*

Attn.:
Commercial Lab Enterprise
270 Swinton Avenue
Bronx, NY 10465

Friday, November 02, 2001

Ref Number: LI0116584

Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES ≥ 0.5µ - < 5µ	# STRUCTURES ≥ 5µ	# STRUCTURES Non-Asbestos	Area Analyzed (mm²)	Analytical Sensitivity (S/cc)	Asbestos Concentration (S/mm²)	Asbestos Concentration (S/cc)
511	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
512	1340.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
513	1350.00	None Detected			0	0.0635	0.0045	<15.75	<0.0045
514	1300.00	None Detected			0	0.0635	0.0047	<15.75	<0.0047
515	1300.00	None Detected			0	0.0635	0.0047	<15.75	<0.0047
516	1300.00	None Detected			0	0.0635	0.0047	<15.75	<0.0047

Earl Tracey
Analyst

Approved Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysis accuracy and precision) is available upon request.

Accreditation for ... NYELAP 11452 ... CT PH-0256

NYC-WTC_000209618

NYC 9/11 Public Portal Document

EMSL Analytical, Inc.

208 Stonehinge Lane
Carle Place, NY 11514
Phone: (516) 997-7251 *Fax: (516) 997-7528*

Attn.:
Commercial Lab Enterprise
270 Swinton Avenue
Bronx, NY 10465

Friday, November 02, 2001

Ref Number: LI0116584

Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES ≥ 0.5µ - < 5µ	# STRUCTURES ≥ 5µ	# STRUCTURES Non-Asbestos	Area Analyzed (mm²)	Analytical Sensitivity (S/cc)	Asbestos Concentration (S/mm²)	Asbestos Concentration (S/cc)
517	1300.00	None Detected			0	0.0635	0.0047	<15.75	<0.0047
518	1280.00	None Detected			0	0.0635	0.0047	<15.75	<0.0047
519	0.00	None Detected			0	0.0635		<15.75	
520	0.00	None Detected			0	0.0635		<15.75	
521	0.00	None Detected			0	0.0635		<15.75	

Earl Tracey
Analyst

Approved Signatory

Disclaimer: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analyst accuracy and precision) is available upon request.

Accredited for NVLAP PLM/TEM #101048-1, NY ELAP 11459, CA ELAP 2349, CT PH-0201

NYC-WTC_000209619