

EMSL Analytical, Inc.208 Stonehinge Lane
Carle Place, NY 11514

Phone: (516) 997-7251 Fax: (516) 997-7528



Attn.:

Taylor Environmental Group, Inc.
Floral Park Corporate Center
130-A Jericho Tpke.
Floral Park, NY 11001

Sunday, September 30, 2001

Ref Number: LI0115210

**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)**

Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES		Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			$\geq 0.5\mu$ - < 5μ	Non- Asbestos $\geq 5\mu$			(S/mm ³)	(S/cc)
1101	1729.00	Chrysotile	1	0	0.0508	0.0044	19.69	0.0044
1102	1719.50	None Detected		0	0.0508	0.0044	<19.69	<0.0044
1103	1710.00	None Detected		0	0.0508	0.0044	<19.69	<0.0044
1104	1681.50	None Detected		0	0.0508	0.0045	<19.69	<0.0045
1105	1691.00	None Detected		0	0.0508	0.0045	<19.69	<0.0045
1001	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041

Michelle McGowan

Analyst


 Approved
Signatory

Disclaimer: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm³. This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

Accredited for NVLAP PLM/TEM #101046-10, NY ELAP 11459, CA ELAP 2339, CT PH-0249

EMSL Analytical, Inc.

208 Stonehenge Lane

Carle Place, NY 11514

Phone: (516) 997-7251 Fax: (516) 997-7528

EMSL

Attn.:

Taylor Environmental Group, Inc.

Floral Park Corporate Center

130-A Jericho Tpke.

Floral Park, NY 11001

Sunday, September 30, 2001

Ref Number: LI0115210


**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)**

Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES		Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			$\geq 0.5\mu$ - < 5μ	Non- Asbestos $\geq 5\mu$			(S/mm ²)	(S/cc)
1002	1803.75	None Detected		0	0.0508	0.0042	<19.69	<0.0042
1003	1852.50	Chrysotile	1	0	0.0508	0.0041	19.69	0.0041
1004	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
1005	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
1006	1852.50	Chrysotile	1	0	0.0508	0.0041	19.69	0.0041
1007	1803.75	None Detected		0	0.0508	0.0042	<19.69	<0.0042


Michelle McGowan

Analyst


 Approved
 Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analyst accuracy and precision) is available upon request.

2


 Accredited for NVLAP PLM/TEM #101048-10, NY ELAP 11489, CA ELAP 2335, CT PH-0249

EMSL Analytical, Inc.

208 Stonehenge Lane

Carle Place, NY 11514

Phone: (516) 997-7251 Fax: (516) 997-7528

EMSL

Attn.:

Taylor Environmental Group, Inc.

Floral Park Corporate Center

130-A Jericho Tpke.

Floral Park, NY 11001

Sunday, September 30, 2001

Ref Number: LI0115210


**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)**

Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES		Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			$\geq 0.5\mu -$ $< 5\mu$	$\geq 5\mu$ Non- Asbestos			(S/mm ²)	(S/cc)
1008	1852.00	None Detected		0	0.0508	0.0041	<19.69	<0.0041
1009	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0901	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0902	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0903	1873.00	None Detected		0	0.0508	0.0040	<19.69	<0.0040
0904	1813.00	None Detected		0	0.0508	0.0042	<19.69	<0.0042

Keith McWilliams

Analyst



Approved
Signatory

Disclaimer: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

3

Accredited for NVLAP PLM/TEM #101048-10 NY ELAP 11469 CA ELAP 2339 CT PH-0249

EMSL Analytical, Inc.

208 Stonehinge Lane
 Carle Place, NY 11514
 Phone: (516) 997-7251 Fax: (516) 997-7528

EMSL

Attn.:

Taylor Environmental Group, Inc.
 Floral Park Corporate Center
 130-A Jericho Tpke.
 Floral Park, NY 11001

Sunday, September 30, 2001

Ref Number: LI0115210

**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
 Performed by EPA 40 CFR Part 763 Final Rule (AHERA)**

Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES		Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			$\geq 0.5\mu$ - < 5μ	Non-Asbestos $\geq 5\mu$			(S/mm ²)	(S/cc)
0905	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0906	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0907	1852.50	Chrysotile	1	0	0.0508	0.0041	19.69	0.0041
0908	1813.00	None Detected		0	0.0508	0.0042	<19.69	<0.0042
0909	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0801	1813.00	None Detected		0	0.0508	0.0042	<19.69	<0.0042

Keith McWilliams

Analyst

Richard A. ...
 Approved
 Signatory

Disclaimer: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysis accuracy and precision) is available upon request.

4

Accredited for: NVLAP PLM/TEM #101048-10 NY ELAP 11489 CA ELAP 2339 CT PH-0249

EMSL Analytical, Inc.208 Stonehenge Lane
Carle Place, NY 11514

Phone: (516) 997-7251 Fax: (516) 997-7528

EMSL

Attn.:

Taylor Environmental Group, Inc.
Floral Park Corporate Center
130-A Jericho Tpke.
Floral Park, NY 11001

Sunday, September 30, 2001

Ref Number: LI0115210


**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)**

Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES		Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			$\geq 0.5\mu$ - < 5μ	Non- Asbestos $\geq 5\mu$			(S/mm ²)	(S/cc)
0802	1813.00	None Detected		0	0.0508	0.0042	<19.69	<0.0042
0803	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0804	1822.25	None Detected		1	0.0508	0.0042	<19.69	<0.0042
0805	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0806	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0807	1852.50	Chrysotile	1	0	0.0508	0.0041	19.69	0.0041

Keith McWilliams

Analyst

Approved
Signatory

Disclaimer: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analyst accuracy and precision) is available upon request.

5

Accredited for NVLAP PLM/TEM #101048-10, NY ELAP 11450, CA ELAP 2339, CT PH-0249

EMSL Analytical, Inc.

208 Stonehenge Lane

Carle Place, NY 11514

Phone: (516) 997-7251 Fax: (516) 997-7528

EMSL

Attn:

Taylor Environmental Group, Inc.

Floral Park Corporate Center

130-A Jericho Tpke.

Floral Park, NY 11001

Sunday, September 30, 2001

Ref Number: LI0115210


**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)**

Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES			Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			≥ 0.5μ - < 5μ	≥ 5μ	Non- Asbestos			(S/mm ²)	(S/cc)
0808	1852.50	Chrysotile	1		0	0.0508	0.0041	19.69	0.0041
0809	1852.50	Chrysotile	1		0	0.0508	0.0041	19.69	0.0041
0701	1803.75	Amosite	2	1	0	0.0508	0.0042	59.06	0.0126
0702	1852.50	Chrysotile	2		0	0.0508	0.0041	39.37	0.0082
0703	1852.50	Chrysotile	2		0	0.0508	0.0041	39.37	0.0082
0704	1803.75	None Detected			0	0.0508	0.0042	<19.69	<0.0042

Michelle McGowan

Analyst



Approved
Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

6

Accredited for NVLAP P/LM/TEM #101048-10 NY ELAP 11469 CA ELAP 2339, CT PH-0249

EMSL Analytical, Inc.208 Stonehenge Lane
Carle Place, NY 11514

Phone: (516) 997-7251 Fax: (516) 997-7528

EMSL

Attn:

Taylor Environmental Group, Inc.

Floral Park Corporate Center

130-A Jericho Tpke.

Floral Park, NY 11001

Sunday, September 30, 2001


Ref Number: LI0115210

**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)****Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS**

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES		Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			$\geq 0.5\mu$ < 5μ	Non- Asbestos $\geq 5\mu$			(S/mm ²)	(S/cc)
0705	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0706	1803.75	None Detected		0	0.0508	0.0042	<19.69	<0.0042
0707	1843.00	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0708	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0709	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0601	1852.50	Chrysotile	1	0	0.0508	0.0041	19.69	0.0041

Keith McWilliams

Analyst

Approved
Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

7

Accredited for NVLAP PLM/TEM #101048-10, NY ELAP 11469, CA ELAP 2339, CT PH-0249

10/01/2001 03:16

5169977528

EMSL LI

PAGE 08

EMSL Analytical, Inc.

208 Stonehenge Lane

Carle Place, NY 11514

Phone: (516) 997-7251 Fax: (516) 997-7528

EMSL

Attn.:

Taylor Environmental Group, Inc.

Floral Park Corporate Center

130-A Jericho Tpke.

Floral Park, NY 11001

Sunday, September 30, 2001

Ref Number: LI0115210


**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)**

Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES		Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			$\geq 0.5\mu$ - < 5μ	$\geq 5\mu$ Non- Asbestos			(S/mm ²)	(S/cc)
0602	1803.75	None Detected		0	0.0508	0.0042	<19.69	<0.0042
0603	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0604	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0605	1803.75	None Detected		0	0.0508	0.0042	<19.69	<0.0042
0606	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0607	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041

Brian Riedener

Analyst


 Approved
 Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

8

Accredited for NVLAP PLM/TEM #101048-10, NY ELAP 11469, CA ELAP 2339, CT PH-0245

CITY_DEP_WORBY_CANNATA_SUBP_000003706

NYC-WTC_000209730

10/01/2001 03:16

5169977528

EMSL LI

PAGE 09

EMSL Analytical, Inc.208 Stonehenge Lane
Carle Place, NY 11514

Phone: (516) 997-7251 Fax: (516) 997-7528



Attn.:

Taylor Environmental Group, Inc.

Floral Park Corporate Center

130-A Jericho Tpke.

Floral Park, NY 11001

Sunday, September 30, 2001

Ref Number: LI0115210

**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)**

Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS

STRUCTURES

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES		Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			$\geq 0.5\mu$ < 5μ	Non- Asbestos $\geq 5\mu$			(S/mm ²)	(S/cc)
0607A	1852.50	None Detected		1	0.0508	0.0041	<19.69	<0.0041
0608	1803.75	None Detected		0	0.0508	0.0042	<19.69	<0.0042
0609	1803.75	None Detected		0	0.0508	0.0042	<19.69	<0.0042
0610	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0611	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0612	1824.00	None Detected		0	0.0508	0.0042	<19.69	<0.0042

Keith McWilliams

Analyst

Approved
Signatory

Disclaimer: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analyst accuracy and precision) is available upon request.

9

Accredited for NVLAP PLM/TEM #101048-10, NY ELAP 11469, CA ELAP 2339, CT PH-0249

10/01/2001 03:16

5169977528

EMSL LI

PAGE 10

EMSL Analytical, Inc.

208 Stonehenge Lane

Carle Place, NY 11514

Phone: (516) 997-7251 Fax: (516) 997-7528

EMSL

Attn.:

Taylor Environmental Group, Inc.

Floral Park Corporate Center

130-A Jericho Tpke.

Floral Park, NY 11001

Sunday, September 30, 2001

Ref Number: LI0115210


Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)

Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES		Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			$\geq 0.5\mu$ < 5μ	$\geq 5\mu$ Non-Asbestos			(S/mm ²)	(S/cc)
0613	1833.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0614	1843.00	Chrysotile	1	0	0.0508	0.0041	19.69	0.0041
0615	1803.75	None Detected		0	0.0508	0.0042	<19.69	<0.0042
0616	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0617	1852.50	None Detected		3	0.0508	0.0041	<19.69	<0.0041
0618	1852.50	None Detected		1	0.0508	0.0041	<19.69	<0.0041


Michelle McGowan

Analyst


Approved
Signatory

Disclaimer: This laboratory is not responsible for data reported in structure/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

10



Revised for NVLAP PLM/TEM #101048-10, NY ELAP 11499, CA ELAP 2339, CT PH-0749

CITY_DEP_WORBY_CANNATA_SUBP_000003708

NYC-WTC_000209732

EMSL Analytical, Inc.

208 Stonehenge Lane

Carle Place, NY 11514

Phone: (516) 997-7251 Fax: (516) 997-7528

EMSL

Attn.:

Taylor Environmental Group, Inc.

Floral Park Corporate Center

130-A Jericho Tpke.

Floral Park, NY 11001

Sunday, September 30, 2001

Ref Number: LI0115210

**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)**

Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES		Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			$\geq 0.5\mu$ - < 5μ	Non-Asbestos $\geq 5\mu$			(S/mm ²)	(S/cc)
0619	1803.75	None Detected		0	0.0508	0.0042	<19.69	<0.0042
0620	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0621	1803.75	None Detected		0	0.0508	0.0042	<19.69	<0.0042
0622	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0623	1843.00	None Detected		0	0.0508	0.0041	<19.69	<0.0041
0624	1852.50	None Detected		0	0.0508	0.0041	<19.69	<0.0041

Michelle McGowan

Analyst


 Approved
Signatory

Disclaimer: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysts' accuracy and precision) is available upon request.

11

Added for NVLAP PLM/TEM #101048-10 NY ELAP 11466, CA ELAP 2339, CT PH-0249

EMSL Analytical, Inc.

208 Stonehinge Lane

Carle Place, NY 11514

Phone: (516) 997-7251 Fax: (516) 997-7528

EMSL

Attn.:

Taylor Environmental Group, Inc.

Floral Park Corporate Center

130-A Jericho Tpke.

Floral Park, NY 11001

Sunday, September 30, 2001

Ref Number: LI0115210


**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM)
Performed by EPA 40 CFR Part 763 Final Rule (AHERA)**

Project: NYC SCA- STUYVESANT HS 6,7,8,9,10,11TH FLOORS

Sample ID	Volume (liters)	Asbestos Type(s)	# STRUCTURES		Area Analyzed (mm ²)	Analytical Sensitivity (S/cc)	Asbestos Concentration	
			$\geq 0.5\mu$ - < 5μ	Non- Asbestos $\geq 5\mu$			(S/mm ²)	(S/cc)
0701 QC	1803.75	Amosite	1	0	0.0508	0.0042	19.69	0.0042
0702 QC	1852.50	Chrysotile	2	0	0.0508	0.0041	39.37	0.0082
0703 QC	1852.50	Chrysotile	2	0	0.0508	0.0041	39.37	0.0082

Daniel Clarke

Analyst


 Approved
Signatory

Disclaimers: This laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This laboratory is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by EMSL. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analytical accuracy and precision) is available upon request.

12

Specified for NVLAP PLM/TEM #101048-10, NY ELAP 11468, CA ELAP 2335, CT PH-0249

