



TEM Asbestos (water) Count Sheetsample area
analyzedClient Name: SCILAB BOSTON, INC.Job #: 101101891Lab Sample #: 01Client Sample #: 0110-372-1Received: 10/30/2001 11:55:14Date Analyzed: 11/01/2001Scope #: 2Volume (liters): 0.02Filter Type / Filter Area: PC 214 mm2Grid Opening Size: 0.01003Area Examined: 0.06018 mm2Magnification: 20,000Accelerating Voltage: 100 KeVAnalysis Performed by: 

Brian Keith


Location	Grid Opening	Fiber	Length μM	Width μM	Fiber Type	Morphology	EDS	Orient	SAED	Photo
C8-C4/4	1	NSD							☐	
C8-E4/4	2	NSD							☐	
C8-F4/4	3	NSD							☐	
C8-G4/4	4	NSD							☐	
C7-C4/4	5	NSD							☐	
C7-E4/4	6	NSD							☐	

NSD: No Asbestos Structures Detected

Comments:

		Concentration (million fibers/liter)	Grid Evaluation
Total Grid Openings:	6		☒ Grid Openings Covered > 50%
Chrysotile Asbestos Structures:	0	<0.178	☒ Intact Grid Opening > 50%
Amphibole Asbestos Structures:	0	<0.178	☒ Undissolved Filter < 10%
Asbestos Structures >=10 microns:	0	<0.178	☒ Folded Replica < 50%
Total Asbestos Structures:	0	<0.178	☒ Filter Loading < 25%
Analytical Sensitivity:		0.178	☒ Particulate Even

CITY_21MC100_002842273

TEM Asbestos (water) Count Sheetsample area
analyzedClient Name: SCILAB BOSTON, INC.Job #: 101101891Volume (liters): 0.02Lab Sample #: 04Filter Type / Filter Area: PC 214 mm²Client Sample #: 0110-372-4Grid Opening Size: 0.01003Received: 10/30/2001 11:55:14Area Examined: 0.06018 mm²Date Analyzed: 11/01/2001Magnification: 20,000Scope #: 2Accelerating Voltage: 100 KeVAnalysis Performed by: 

Brian Keith



Location	Grid Opening	Fiber	Length μ M	Width μ M	Fiber Type	Morphology	EDS	Orient.	SAED	Photo
D8-C4/4	1	NSD							☐	
D8-E4/4	2	NSD							☐	
D8-F4/4	3	NSD							☐	
D9-C4/4	4	NSD							☐	
D9-E4/4	5	NSD							☐	
D9-F4/4	6	NSD							☐	

NSD: No Asbestos Structures Detected

Comments:

		Concentration (million fibers/liter)	Grid Evaluation
Total Grid Openings:	6		☒ Grid Openings Covered > 50%
Chrysotile Asbestos Structures:	0	<0.178	☒ Intact Grid Opening > 50%
Amphibole Asbestos Structures:	0	<0.178	☒ Undissolved Filter < 10%
Asbestos Structures $\geq$ 10 microns:	0	<0.178	☒ Folded Replica < 50%
Total Asbestos Structures:	0	<0.178	☒ Filter Loading < 25%
Analytical Sensitivity:		0.178	☒ Particulate Even

CITY_21MC100_002842276

TEM Asbestos (water) Count Sheetsample area
analyzedClient Name: SCILAB BOSTON, INC.Job #: 101101891Volume (liters): 0.02Lab Sample #: 05Filter Type / Filter Area: PC 214 mm²Client Sample #: 0110-372-5Grid Opening Size: 0.01003Received: 10/30/2001 11:55:14Area Examined: 0.06018 mm²Date Analyzed: 11/01/2001Magnification: 20,000Scope #: 2Accelerating Voltage: 100 KeVAnalysis Performed by: 

Brian Keith


Location	Grid Opening	Fiber	Length μ M	Width μ M	Fiber Type	Morphology	EDS	Orient.	SAED	Photo
E6-C4/4	1	NSD							☐	
E6-E4/4	2	NSD							☐	
E6-F4/4	3	NSD							☐	
D10-C4/4	4	NSD							☐	
D10-E4/4	5	NSD							☐	
D10-F4/4	6	NSD							☐	

NSD: No Asbestos Structures Detected

Comments:

		Concentration	Grid Evaluation
Total Grid Openings:	6	(million fibers/liter)	☒ Grid Openings Covered > 50%
Chrysotile Asbestos Structures:	0	<0.178	☒ Intact Grid Opening > 50%
Amphibole Asbestos Structures:	0	<0.178	☒ Undissolved Filter < 10%
Asbestos Structures $\geq$ 10 microns:	0	<0.178	☒ Folded Replica < 50%
Total Asbestos Structures:	0	<0.178	☒ Filter Loading < 25%
Analytical Sensitivity:		0.178	☒ Particulate Even

CITY_21MC100_002842277

TEM Asbestos (water) Count Sheetsample area
analyzedClient Name: SCILAB BOSTON, INC.Job #: 101101891Volume (liters): 0.02Lab Sample #: 06Filter Type / Filter Area: PC 214 mm²Client Sample #: 0110-372-6Grid Opening Size: 0.01003Received: 10/30/2001 11:55:14Area Examined: 0.06018 mm²Date Analyzed: 11/01/2001Magnification: 20,000Scope #: 2Accelerating Voltage: 100 KeVAnalysis Performed by: Brian Keith

Brian Keith

Location	Grid Opening	Fiber	Length μ M	Width μ M	Fiber Type	Morphology	EDS	Orient.	SAED	Photo
E7-C4/4	1	NSD							☐	
E7-E4/4	2	NSD							☐	
E7-F4/4	3	NSD							☐	
E8-C4/4	4	NSD							☐	
E8-E4/4	5	NSD							☐	
E8-F4/4	6	NSD							☐	

NSD: No Asbestos Structures Detected

Comments:

		Concentration	Grid Evaluation
Total Grid Openings:	6	(million fibers/liter)	☒ Grid Openings Covered > 50%
Chrysotile Asbestos Structures:	0	<0.178	☒ Intact Grid Opening > 50%
Amphibole Asbestos Structures:	0	<0.178	☒ Undissolved Filter < 10%
Asbestos Structures $\geq$ 10 microns:	0	<0.178	☒ Folded Replica < 50%
Total Asbestos Structures:	0	<0.178	☒ Filter Loading < 25%
Analytical Sensitivity:		0.178	☒ Particulate Even

CITY_21MC100_002842278