NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	ATC	**Volume (liters):**	1300
Lab Group #:	5994	**Filter Type / Filter Area:**	MCE 385 mm²
Lab Sample #:	16	**Grid Opening Size:**	0.01124
Client Sample #:	16	**Area Examined:**	0 06744 mm²
GRID BOX:	SFF	**Magnification:**	20,000
GRID ID:	FAB	**Accelerating Voltage:**	100 KeV
# Grid Openings:	6	**Analysis Date:**	11/29/01

Analysis Performed By: MARK PEYSAKHOV

GO No	GO Coord	Structure No	Structure Type F	B	C	M	Length >=05 µ <5µ	>=5µ	ID of Structure N or √ Chrys	Amphibole	Nonasb	Ambig/ND	EDS (Nonadb Elements)
1	D1	NSD											
2	C1	—											
3	A1	—											
4	JJ	—											
5	AA	—											
6	XX	—											

Struct = > 0 5µ .> 5 1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart

C = Cluster (>3 Intersections)

M = Matnx

NSD = No Structures Detected

Grid acceptance

<50% undissolved filter, overlapped or folded <10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES

Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

CITY_DEP_WORBY_CANNATA_SUBP_000002801

NYC-WTC_000210647