


TEM Asbestos (air) Count Sheetsample area
analyzedClient Name: ACS Environmental Services, Inc.Job #: 201103845Volume (liters): 1 280.0Lab Sample #: 02Filter Type / Filter Area: MCE 385 mm²Client Sample #: 0225.32.2Grid Opening Size: 0.01000Received: 10/30/2001

13:51:10

Area Examined: 0.07000 mm²Date Analyzed: 10/30/2001Magnification: 20,000Scope #:Accelerating Voltage: 100 KeVAnalysis Performed by: RPly
Roman Peysakhov

Location	Grid Opening	Fiber	Length μM	Width μM	Fiber Type	Morphology	EDS	Orient.	SAED	Photo
A3-4/1C	1	NSD							☐	
A3-4/1E	2	NSD							☐	
A3-4/1F	3	NSD							☐	
A3-4/1G	4	NSD							☐	
A4-4/1C	5	NSD							☐	
A4-4/1E	6	NSD							☐	
A4-4/1F	7	NSD							☐	

NSD: No Asbestos Structures Detected

Comments

	Structure Density (str/mm ²)	Concentration (str/cc air)	<u>Grid Evaluation</u>
Total Grid Openings: 7			☒ Grid Openings Covered > 50%
Chrysotile Asbestos Structures: 0	<14.3	<0.0043	☒ Intact Grid Opening > 50%
Amphibole Asbestos Structures: 0	<14.3	<0.0043	☒ Undissolved Filter < 10%
Asbestos Structures ≥ 5 microns: 0	<14.3	<0.0043	☒ Folded Replica < 50%
Total Non-Asbestos Structures:			☒ Filter Loading < 10%
Total Asbestos Structures: 0	<14.3	<0.0043	☒ Particulate Even
Analytical Sensitivity:	14.3	0.0043	

CITY_21MC100_002850356