

**ATC Associates**

104 East 25 Street, New York, NY 10010

Phone (212) 353-8280 Fax (212) 353-8306

Attn: Mr. Joshua Sarent  
ALC Environmental

115 W 30th Street

New York NY 10001

Fax: (212) 279-6747 Phone: (212) 279-1001

Project: Gateway Plaza Management

375 South End Ave , Bldg 100 New York , NY

Received: 1/21/02 5 00 00 PM

ATC Group #: 6399

Analysis Date: 1/22/02

**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM) Performed  
by EPA 40 CFR Part 763 Final Rule (AHERA)**

| Sample         | Volume  | Area<br>Analyzed | Non<br>Asb | Asbestos<br>Type(s) | Structures<br>0.5µ-5 µ |   | Analytical<br>Sensitivity<br>(S/cc) | Asbestos<br>Concentration<br>(S/mm) | (S/cc)  | Notes |
|----------------|---------|------------------|------------|---------------------|------------------------|---|-------------------------------------|-------------------------------------|---------|-------|
| D-1<br>6399 -1 | 1800.00 | 0.0450           | 0          | None Detected       | 0                      | 0 | 0.0048                              | <22.24                              | <0.0048 |       |
| D-2<br>6399 -2 | 1800.00 | 0.0450           | 0          | None Detected       | 0                      | 0 | 0.0048                              | <22.24                              | <0.0048 |       |
| D-3<br>6399 -3 | 1800.00 | 0.0450           | 0          | None Detected       | 0                      | 0 | 0.0048                              | <22.24                              | <0.0048 |       |
| D-4<br>6399 -4 | 1800.00 | 0.0450           | 0          | None Detected       | 0                      | 0 | 0.0048                              | <22.24                              | <0.0048 |       |
| D-5<br>6399 -5 | 1800.00 | 0.0450           | 0          | None Detected       | 0                      | 0 | 0.0048                              | <22.24                              | <0.0048 |       |
| D-6<br>6399 -6 | 1800.00 | 0.0450           | 0          | None Detected       | 0                      | 0 | 0.0048                              | <22.24                              | <0.0048 |       |
| D-7<br>6399 -7 | 1800.00 | 0.0450           | 0          | None Detected       | 0                      | 0 | 0.0048                              | <22.24                              | <0.0048 |       |
| D-8<br>6399 -8 | 1800.00 | 0.0450           | 0          | None Detected       | 0                      | 0 | 0.0048                              | <22.24                              | <0.0048 |       |

MARK PEYSAKHOV

Analyzed by:

MICHAEL A. GORYCKI, Ph.D.

Approved by

Disclaimer: The laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This lab is only responsible for data reported in structures/mm. This report may not be reproduced, except in full, without written approval by ATC Associates Inc. This report may not be used to claim product endorsement by NVLAP or any other agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysis accuracy and precision) is available upon request.

Accredited for NVLAP PLW-TEM #101187-0 NY State ELAP #10879

Attn: Mr. Joshua Saret  
 ALC Environmental  
 115 W. 30th Street  
 New York NY 10001  
 Fax: (212) 279-6747 Phone: (212) 279-1001  
 Project: Gateway Plaza Management  
 375 South End Ave., Bldg. 100 New York, NY


Received: 1/21/02 5:00:00 PM  
 ATC Group #: 6399  
 Analysis Date: 1/22/02

**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM) Performed  
 by EPA 40 CFR Part 763 Final Rule (AHERA)**

| Sample           | Volume  | Area<br>Analyzed | Non<br>Ash | Asbestos<br>Type(s) | Structures<br>0.5µ-5µ | 5µ | Analytical<br>Sensitivity<br>(S/cc) | Asbestos<br>Concentration<br>(S/mm <sup>3</sup> ) | (S/cc)   | Notes |
|------------------|---------|------------------|------------|---------------------|-----------------------|----|-------------------------------------|---------------------------------------------------|----------|-------|
| D-9<br>6399 -9   | 1800.00 | 0.0450           | 0          | None Detected       | 0                     | 0  | 0.0048                              | < 22.24                                           | < 0.0048 |       |
| D-10<br>6399 -10 | 1500.00 | 0.0450           | 0          | None Detected       | 0                     | 0  | 0.0048                              | < 22.24                                           | < 0.0048 |       |

MARK PEYSAKHOV

Analyzed by



MICHAEL A. GORYCKI, Ph.D.

Approved by

Disclaimer: The laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. The lab is only responsible for data reported in structures/mm<sup>3</sup>. This report may not be reproduced, except in full, without written approval by ATC Associates Inc. This report may not be used to claim product endorsement by NVLAP or any other agency of the U.S. Government. This report relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysis accuracy and precision) is available upon request.

Accredited for NVLAP P.M/TEM #101187-C NY State ELAP #10879