



INDUSTRIAL HYGIENE

ENVIRONMENTAL TESTING

• EPA NVLAP 101252-0
• AIHA ACCREDITATION NO. 100439
• NC DENR 598

• NY DOH/NELAC 10903
• PA DEP 06-353

• NJ DEP PA020
• CT DPH PH-0238

**Analytical Results
for**

ATC Associates

WorkOrder: R02031091

Client Reference: 19125-0290

Analyte	Concentration			Limit of Detection (ug)	Test Method	Date Analyzed /Analyst	
	(ug)	(mg/m³)	(ppm)				
Client ID: ATC-0322-M894-S1 <i>Main Lobby</i>	Lab ID: -01A	Date Sampled: 3/22/02		Matrix: Filter		Air Vol.(L): 1020	
Total Particulate Matter	<50.0	<0.0490	--	50	N0500	04/02/2002	ADM
Crystalline Silica	<10.0	<0.00980	--	10	Crystalline Silica	04/02/2002	ADM
Client ID: ATC-0322-M894-S2 <i>5th floor</i>	Lab ID: -02A	Date Sampled: 3/22/02		Matrix: Filter		Air Vol.(L): 1020	
Total Particulate Matter	<50.0	<0.0490	--	50	N0500	04/02/2002	ADM
Crystalline Silica	<10.0	<0.00980	--	10	Crystalline Silica	04/02/2002	ADM
Client ID: ATC-0322-M894-S3 <i>10th floor</i>	Lab ID: -03A	Date Sampled: 3/22/02		Matrix: Filter		Air Vol.(L): 1020	
Total Particulate Matter	<50.0	<0.0490	--	50	N0500	04/02/2002	ADM
Crystalline Silica	<10.0	<0.00980	--	10	Crystalline Silica	04/02/2002	ADM
Client ID: ATC-0322-M894-S4 <i>Blank</i>	Lab ID: -04A	Date Sampled: 3/22/02		Matrix: Filter		Air Vol.(L): NA	
Total Particulate Matter	<50.0	--	--	50	N0500	04/02/2002	ADM
Crystalline Silica	<10.0	--	--	10	Crystalline Silica	04/02/2002	ADM

(a) Analysis indicates possible breakthrough; back section result is greater than 0% of the front section result.

General Notes:

<: Less than the indicated limit of detection (LOD).

--: Information not available or not applicable.

Back sections were checked and showed no significant breakthrough.