NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	**Volume (lit**	1920
Lab Group #:	8212	**Filter Type / Filter Ar**	MCE 385 mm²
Lab Sample #:	12	**Grid Opening S**	0.01124
Client Sample #:	12	**Area Examin**	0.04496 mm²
GRID BOX:	1257	**Magnificati**	20,000
GRID ID:	SCM	**Accelerating Volt**	100 KeV
# Grid Openings:	4	**Analysis D**	6/22/02

Analysis Performed By: MARK PEYSAKHOV

GO. No	GO. Coord	Structure No	Structure Type F	B	C	M	Length >=05 µ <5µ	Length >=5µ	ID of Structure Chrys	Amphibole	Nona	√ Amphig/ND	EDS (Nonadb Elements)
SC	1 AQ	NSD											
—	2 JQ	—											
SC	3 KG	—											
	4 KG	—											

Struct = > 0.5µ > 5:1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber
B = Bundle (<3 fibers <1 fiber diameter apart
C = Cluster (>3 Intersections)
M = Matrix
NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

	B	C	M	>5µ
Chrysotile				
Amphibole				
Ambiguous				
Non-Asbestos				
Non-Identified				

NYC-WTC_000214848