NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	**Volume (li**	1920
Lab Group #:	8212	**Filter Type / Filter A**	MCE 385 mm²
Lab Sample #:	15	**Grid Opening**	0.01124
Client Sample #:	15	**Area Exami**	0.04496 mm²
GRID BOX:	1257	**Magnifica**	20,000
GRID ID:	6470	**Accelerating Vo**	100 KeV
# Grid Openings:	4	**Analysis**	6/22/02

Analysis Performed By: MARK PEYSAKHOV

GO. No	GO. Coord	Structure No	Structure Type F	B	C	M	Length >=05 µ <5µ	Length >=5µ	ID of Structure Chrys	Amphibole	n.c.	√ Ambig/ND	EDS (Nonadb Elements)
1	C 4	NSD											
2	S E	1											
3	R I	1											
4	E R I	1											

Struct = > 0.5µ > 5:1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart

C = Cluster (>3 Intersections)

M = Matrix

NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded

<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES

Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Chry	
Ampl	
Ambi	
Non-Asb	
Non-Identified	

ıls

B	C	M	>5µ

NYC-WTC_000214851