

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	Volume (liters):	1928
Lab Group #:	7233	Filter Type / Filter Area:	MCE 385 mm ²
Lab Sample #:	20	Grid Opening Size:	0.01124
Client Sample #:	20	Area Examined:	0.04496 mm ²
GRID BOX:	9821	Magnification:	20,000
GRID ID:	1AC	Accelerating Voltage:	100 KeV
# Grid Openings:	4	Analysis Date:	3/27/02

Analysis Performed By: RP
ROMAN PEYSAKHOV

GO.	No	Coord	Structure No	Structure Type				Length		ID of Structure: N or ✓				EDS (Nonadb Elements)
				F	B	C	M	>=05 μ <5 μ	>=5 μ	Chrys	Amphibole	Nonasb	Ambig/ND	
1A	1	I2	NSD											
	2	P2	4											
	3	S2	4											
	4	B2	4											
1C														

Struct = > 0.5 μ > 5:1 Aspect Ratio

✓ = Positive diffraction pattern properties conforming to laboratory identification criteria observed

✓ = EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart)

C = Cluster (>3 intersections)

M = Matrix

NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded

<10% particulate coverage >50% coverage at least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES

Amosite

Crocidolite

Anthophyllite

Tremolite

Actinolite

Totals					
	F	B	C	M	>5 μ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					