

ENVIRONMENTAL CONSULTANTS, INC.
 200 Park Avenue South, Suite 1106, New York, New York 10003
 Phone: (212) 420-8119 Fax: (212) 420-6092

WSA Management
 100 Ring Road West
 Garden City, NY 11530

ELAP Lab Code: 11029
 NIOSH and AIHA Lab Code: 100273
 NVLAP Lab Code: 101953

Project No.: 01-2214
 Analyte: Airborn Particulate
 Date Collected: 11/16/01
 Date Received: 11/16/01
 Date Analyzed: 11/16/01

Site: 80 John Street, New York, NY

Summary of Analytical Results

SAMPLE # LAB #	LOCATION / DESCRIPTION	VOLUME	PLOD * (mg / m3)	CONCENTRATION (mg / m3)
1 1116-01	Ambient/ 8th Floor/ Apartment 8-I/ South Section/ Bedroom Door (On West)	120		0.58
2 1116-02	Ambient/ 8th Floor/ Apartment 8-I/ North Section/ Living Room/ North West Door	120		0.42
3 1116-03	Ambient/ 8th Floor/ Corridor/ North End of Corridor with North Wall	120		0.92
4 1116-04	Ambient/ 8th Floor/ Corridor/ South End of Corridor with South Wall	120		0.75
5 1116-05	Ambient/ 22nd Floor/ Apartment 22-A/ South Side Living Room/ In the Middle of South Wall	120		0.50
6 1116-06	Ambient/ 22nd Floor/ Apartment 22-A/ North East Room/ On West Side Near Door	120		0.08
7 1116-07	FIELD BLANK			<0.01/ Filter

Sample Analysis by:

Analytical Balance


* PLOD = Performance limit of detection, which is 0.2 mg/ air volume (l).

Method of Sample Preparation and Analysis

All samples were prepared and analyzed in accordance with the National Institute of Occupational Safety and Health (NIOSH) Method 0500 02/15/84.

Instrumentation:

Denver A-200DS, with resolution at 0.01 mg.



Gulam Machhiwala
 Laboratory Director

OSHA Permissible Exposure Limit (PEL) for total Nuisance Dust is 5 mg/m3

ENVIRONMENTAL CONSULTANTS, INC.
 200 Park Avenue South, Suite 1106, New York, New York 10003
 Phone: (212) 420-8119 Fax: (212) 420-6092

WSA Management
 100 Ring Road West
 Garden City, NY 11530

ELAP Lab Code: 11029
 NIOSH and AIHA Lab Code: 100273
 NVLAP Lab Code: 101953

Project No.: 01-2214
 Analyte: Airborn Particulate
 Date Collected: 11/16/01
 Date Received: 11/16/01
 Date Analyzed: 11/16/01

Site: 80 John Street, New York, NY

Summary of Analytical Results

SAMPLE #	LOCATION / DESCRIPTION	VOLUME	PLOD *	CONCENTRATION
LAB #			(mg / m3)	(mg / m3)
8	FIELD BLANK			<0.01/ Filter
1116-08				

* PLOD = Performance limit of detection, which is 0.2 mg/ air volume (l).

Sample Analysis by:


Analytical Balance

Method of Sample Preparation and Analysis

All samples were prepared and analyzed in accordance with the National Institute of Occupational Safety and Health (NIOSH) Method 0500 02/15/84.

Instrumentation:

Denver A-200DS, with resolution at 0.01 mg.



Gulam Machhiwala
 Laboratory Director

OSHA Permissible Exposure Limit (PEL) for total Nuisance Dust is 5 mg/m3