



ANALYTICAL SERVICES

INDUSTRIAL HYGIENE

• EPA NMLAP 101262-0
• AIHA ACCREDITATION NO. 100439
• NC DENR 599

ENVIRONMENTAL TESTING

• NY DOH NELAC 10903
• PA DEP 06-353
• NJ DEP PA020
• CT DPH PH-0238

Analytical Results
for

ATC Associates

WorkOrder: R02040887

Client Reference: 19125-0297, Mercury

Analyte	Concentration			Limit of Detection (ug)	Test Method	Date Analyzed /Analyst
	(ug)	(mg/m³)	(ppm)			
Client ID: Hg1 ATC-0418-89M	Lab ID: -01A	Date Sampled: 4/18/02		Matrix: HTBE		Air Vol.(L): 120
Mercury	0.0310	0.000258	--	0.01	N6009M	04/25/2002 JDC
Client ID: Hg2 ATC-0418-89M	Lab ID: -02A	Date Sampled: 4/18/02		Matrix: HTBE		Air Vol.(L): 120
Mercury	0.0270	0.000225	--	0.01	N6009M	04/25/2002 JDC
Client ID: Hg3 ATC-0418-89M	Lab ID: -03A	Date Sampled: 4/18/02		Matrix: HTBE		Air Vol.(L): 120
Mercury	0.0290	0.000242	--	0.01	N6009M	04/25/2002 JDC
Client ID: Hg4 ATC-0418-89M	Lab ID: -04A	Date Sampled: 4/18/02		Matrix: HTBE		Air Vol.(L): 120
Mercury	0.0290	0.000242	--	0.01	N6009M	04/25/2002 JDC
Client ID: Hg5 ATC-0418-89M	Lab ID: -05A	Date Sampled: 4/18/02		Matrix: HTBE		Air Vol.(L): NA
Mercury	0.0350	--	--	0.01	N6009M	04/25/2002 JDC

(a) Analysis indicates possible breakthrough; back section result is greater than 0% of the front section result.

General Notes:

<: Less than the indicated limit of detection (LOD).

--: Information not available or not applicable.

Back sections were checked and showed no significant breakthrough.