



INDUSTRIAL HYGIENE

ENVIRONMENTAL TESTING

• EPA NVLAP 101262-0
• AIHA ACCREDITATION NO. 100439
• NC DENR 599

• NY DOH NELAC 10903
• PA DEP 06-353

• NJ DEP PA020
• CT DPH PH-0238

**Analytical Results
for**

ATC Associates

WorkOrder: R02040881

Client Reference: 19125-0297, Metals

Analyte	Concentration			Limit of Detection (ug)	Test Method	Date Analyzed /Analyst
	(ug)	(mg/m³)	(ppm)			

Client ID: ATC-0418-89M-M1	Lab ID: -01A	Date Sampled: 4/18/02		Matrix: Filter		Air Vol.(L): 1200	
Aluminum	<1.00	<0.000833	--	1	N7300 MCE	04/29/2002	AJB
Cadmium	<0.200	<0.000167	--	0.2	N7300 MCE	04/29/2002	AJB
Chromium	0.446	0.000371	--	0.2	N7300 MCE	04/29/2002	AJB
Cobalt	<0.100	<0.0000833	--	0.1	N7300 MCE	04/29/2002	AJB
Copper	<0.200	<0.000167	--	0.2	N7300 MCE	04/29/2002	AJB
Iron	<2.00	<0.00167	--	2	N7300 MCE	04/29/2002	AJB
Lead	0.374	0.000311	--	0.3	N7300 MCE	04/29/2002	AJB
Magnesium	<1.30	<0.00108	--	1.3	N7300 MCE	04/29/2002	AJB
Manganese	<0.300	<0.000250	--	0.3	N7300 MCE	04/29/2002	AJB
Molybdenum	<1.00	<0.000833	--	1	N7300 MCE	04/29/2002	AJB
Nickel	<0.200	<0.000167	--	0.2	N7300 MCE	04/29/2002	AJB
Zinc	<0.500	<0.000417	--	0.5	N7300 MCE	04/29/2002	AJB

Client ID: ATC-0418-89M-M2	Lab ID: -02A	Date Sampled: 4/18/02		Matrix: Filter		Air Vol.(L): 1200	
Aluminum	<1.00	<0.000833	--	1	N7300 MCE	04/30/2002	AJB
Cadmium	<0.200	<0.000167	--	0.2	N7300 MCE	04/30/2002	AJB
Chromium	0.576	0.000480	--	0.2	N7300 MCE	04/30/2002	AJB
Cobalt	<0.100	<0.0000833	--	0.1	N7300 MCE	04/30/2002	AJB
Copper	<0.200	<0.000167	--	0.2	N7300 MCE	04/30/2002	AJB
Iron	<2.00	<0.00167	--	2	N7300 MCE	04/30/2002	AJB
Lead	<0.300	<0.000250	--	0.3	N7300 MCE	04/30/2002	AJB
Magnesium	<1.30	<0.00108	--	1.3	N7300 MCE	04/30/2002	AJB
Manganese	<0.300	<0.000250	--	0.3	N7300 MCE	04/30/2002	AJB
Molybdenum	<1.00	<0.000833	--	1	N7300 MCE	04/30/2002	AJB
Nickel	<0.200	<0.000167	--	0.2	N7300 MCE	04/30/2002	AJB
Zinc	<0.500	<0.000417	--	0.5	N7300 MCE	04/30/2002	AJB



INDUSTRIAL HYGIENE

ENVIRONMENTAL TESTING

• EPANVLAP 101262-0
• AIHA ACCREDITATION NO. 100439
• NC DENR 599

• NY DOH/NELAC 10903
• PA DEP 06-353

• NJ DEP PA020
• CT DPH PH-0238

Analytical Results
for

ATC Associates

WorkOrder: R02040881

Client Reference: 19125-0297, Metals

Analyte	Concentration			Limit of Detection (ug)	Test Method	Date Analyzed /Analyst	
	(ug)	(mg/m ³)	(ppm)				

Client ID: ATC-0418-89M-M3	Lab ID: -03A	Date Sampled: 4/18/02		Matrix: Filter		Air Vol.(L): 1200	
Aluminum	<1.00	<0.000833	--	1	N7300 MCE	04/30/2002	AJB
Cadmium	<0.200	<0.000167	--	0.2	N7300 MCE	04/30/2002	AJB
Chromium	0.499	0.000416	--	0.2	N7300 MCE	04/30/2002	AJB
Cobalt	<0.100	<0.0000833	--	0.1	N7300 MCE	04/30/2002	AJB
Copper	<0.200	<0.000167	--	0.2	N7300 MCE	04/30/2002	AJB
Iron	<2.00	<0.00167	--	2	N7300 MCE	04/30/2002	AJB
Lead	<0.300	<0.000250	--	0.3	N7300 MCE	04/30/2002	AJB
Magnesium	<1.30	<0.00108	--	1.3	N7300 MCE	04/30/2002	AJB
Manganese	<0.300	<0.000250	--	0.3	N7300 MCE	04/30/2002	AJB
Molybdenum	<1.00	<0.000833	--	1	N7300 MCE	04/30/2002	AJB
Nickel	<0.200	<0.000167	--	0.2	N7300 MCE	04/30/2002	AJB
Zinc	<0.500	<0.000417	--	0.5	N7300 MCE	04/30/2002	AJB

Client ID: ATC-0418-89M-M4	Lab ID: -04A	Date Sampled: 4/18/02		Matrix: Filter		Air Vol.(L): 1200	
Aluminum	<1.00	<0.000833	--	1	N7300 MCE	04/30/2002	AJB
Cadmium	<0.200	<0.000167	--	0.2	N7300 MCE	04/30/2002	AJB
Chromium	0.632	0.000527	--	0.2	N7300 MCE	04/30/2002	AJB
Cobalt	<0.100	<0.0000833	--	0.1	N7300 MCE	04/30/2002	AJB
Copper	<0.200	<0.000167	--	0.2	N7300 MCE	04/30/2002	AJB
Iron	<2.00	<0.00167	--	2	N7300 MCE	04/30/2002	AJB
Lead	<0.300	<0.000250	--	0.3	N7300 MCE	04/30/2002	AJB
Magnesium	<1.30	<0.00108	--	1.3	N7300 MCE	04/30/2002	AJB
Manganese	<0.300	<0.000250	--	0.3	N7300 MCE	04/30/2002	AJB
Molybdenum	<1.00	<0.000833	--	1	N7300 MCE	04/30/2002	AJB
Nickel	<0.200	<0.000167	--	0.2	N7300 MCE	04/30/2002	AJB
Zinc	<0.500	<0.000417	--	0.5	N7300 MCE	04/30/2002	AJB



ANALYTICAL SERVICES

INDUSTRIAL HYGIENE

ENVIRONMENTAL TESTING

• EPA NVLAP 101262-0

• AIHA ACCREDITATION NO. 100439

• NC DENR 599

• NY DOH/NELAC 10903

• PA DEP 08-353

• NJ DEP PA020

• CT DPH PH-0238

Analytical Results
for

ATC Associates

WorkOrder: R02040881

Client Reference: 19125-0297, Metals

Analyte	Concentration			Limit of Detection (ug)	Test Method	Date Analyzed /Analyst
	(ug)	(mg/m ³)	(ppm)			

Client ID: ATC-0418-89M-M5	Lab ID: -05A	Date Sampled: 4/18/02		Matrix: Filter		Air Vol.(L): NA	
Aluminum	<1.00	--	--	1	N7300 MCE	04/30/2002	AJB
Cadmium	<0.200	--	--	0.2	N7300 MCE	04/30/2002	AJB
Chromium	0.435	--	--	0.2	N7300 MCE	04/30/2002	AJB
Cobalt	<0.100	--	--	0.1	N7300 MCE	04/30/2002	AJB
Copper	<0.200	--	--	0.2	N7300 MCE	04/30/2002	AJB
Iron	<2.00	--	--	2	N7300 MCE	04/30/2002	AJB
Lead	<0.300	--	--	0.3	N7300 MCE	04/30/2002	AJB
Magnesium	<1.30	--	--	1.3	N7300 MCE	04/30/2002	AJB
Manganese	<0.300	--	--	0.3	N7300 MCE	04/30/2002	AJB
Molybdenum	<1.00	--	--	1	N7300 MCE	04/30/2002	AJB
Nickel	<0.200	--	--	0.2	N7300 MCE	04/30/2002	AJB
Zinc	<0.500	--	--	0.5	N7300 MCE	04/30/2002	AJB

Client ID: ATC-0418-89M-M6	Lab ID: -06A	Date Sampled: 4/18/02		Matrix: Filter		Air Vol.(L): NA	
Aluminum	<1.00	--	--	1	N7300 MCE	04/30/2002	AJB
Cadmium	<0.200	--	--	0.2	N7300 MCE	04/30/2002	AJB
Chromium	0.551	--	--	0.2	N7300 MCE	04/30/2002	AJB
Cobalt	<0.100	--	--	0.1	N7300 MCE	04/30/2002	AJB
Copper	<0.200	--	--	0.2	N7300 MCE	04/30/2002	AJB
Iron	<2.00	--	--	2	N7300 MCE	04/30/2002	AJB
Lead	<0.300	--	--	0.3	N7300 MCE	04/30/2002	AJB
Magnesium	<1.30	--	--	1.3	N7300 MCE	04/30/2002	AJB
Manganese	<0.300	--	--	0.3	N7300 MCE	04/30/2002	AJB
Molybdenum	<1.00	--	--	1	N7300 MCE	04/30/2002	AJB
Nickel	<0.200	--	--	0.2	N7300 MCE	04/30/2002	AJB
Zinc	<0.500	--	--	0.5	N7300 MCE	04/30/2002	AJB

(a) Analysis indicates possible breakthrough; back section result is greater than 0% of the front section result.

General Notes:

<: Less than the indicated limit of detection (LOD).

--: Information not available or not applicable.

Back sections were checked and showed no significant breakthrough.