NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	Volume (liters):	1920
Lab Group #:	6479	Filter Type / Filter Area:	MCE 385 mm²
Lab Sample #:	6	Grid Opening Size:	0.01124
Client Sample #:	6	Area Examined:	0.04496 mm²
GRID BOX:	772	Magnification:	20,000
GRID ID:	2030	Accelerating Voltage:	100 KeV
# Grid Openings:	4	Analysis Date:	1/27/02

Analysis Performed By: RP
ROMAN PEYSAKHOV

GO	Structure	Structure Type				Length		ID of Structure N or √				EDS
No Coord	No	F	B	C	M	>=05 µ <5µ	>=5µ	Chrys	Amphibole	Nonasb	Ambig/ND	(Nonadb Elements)
1 I 2	NSD											
2 P 2	4											
3 Q 2	11											
4 J 2	11											

20 / 30

Struct = > 0.5µ > 5:1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber
B = Bundle (<3 fibers <1 fiber diameter apart
C = Cluster (>3 intersections)
M = Matrix
NSD = No Structures Detected

Grid acceptance:
<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

NYC-WTC_000223401