NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	**Volume (liters):**	1920
Lab Group #:	7371	**Filter Type / Filter Area:**	MCE 385 mm²
Lab Sample #:	2	**Grid Opening Size:**	0.01124
Client Sample #:	02	**Area Examined:**	0.04496 mm²
GRID BOX:	1022	**Magnification:**	20,000
GRID ID:	1022-D	**Accelerating Voltage:**	100 KeV
# Grid Openings:	4	**Analysis Date:**	4.10.02

Analysis Performed By: _____

MARK PEYSAKHOV

GO		Structure No	Structure Type				Length		ID of Structure N or V				EDS (Nonasb Elements)
No	Coord		F	B	C	M	>=05 µ <5µ	>=5µ	Chrys	Amphibole	Nonasb	Ambig NC	
1022-1	1 H1	NSD											
	2 H7	"											
	3 D2	"											
1022	4 D7	"											

Struct = > 0.5µ > 5:1 Aspect Ratio

V = Positive diffraction pattern properties conforming to laboratory identification criteria observed

V = EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart

C = Cluster (>3 intersections)

M = Matrix

NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES

Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

NYC-WTC_000223672