NYC 9/11 Public Portal Document



PSC ANALYTICAL SERVICES

INDUSTRIAL HYGIENE ENVIRONMENTAL TESTING

•EPA/NVLAP 101262-0 •NY DOH/NELAC 10903 •NJ DEP PA020
•AIHA ACCREDITATION NO. 100439 • PA DEP 06-353 •CT DPH PH-0238
•NC DENR 599

**Analytical Results**
**for**
**ATC Associates**

**WorkOrder:** R02060664
**Client Reference:** 19125-0290, Mercury

| Analyte | Concentration (ug) | Concentration (mg/m³) | Concentration (ppm) | Limit of Detection (ug) | Test Method | Date Analyzed /Analyst |
|---|---|---|---|---|---|---|
| **Client ID:** ATC-0614-M894-Hg1 | **Lab ID:** -01A | **Date Sampled:** 6/14/2002 | | **Matrix:** HTBE | | **Air Vol.(L):** 120 |
| Mercury | 0.0270 | 0.000225 | -- | 0.01 | N6009M | 06/21/2002 JDC |
| **Client ID:** ATC-0614-M894-Hg2 | **Lab ID:** -02A | **Date Sampled:** 6/14/2002 | | **Matrix:** HTBE | | **Air Vol.(L):** 120 |
| Mercury | 0.0254 | 0.000212 | -- | 0.01 | N6009M | 06/21/2002 JDC |
| **Client ID:** ATC-0614-M894-Hg3 | **Lab ID:** -03A | **Date Sampled:** 6/14/2002 | | **Matrix:** HTBE | | **Air Vol.(L):** 120 |
| Mercury | 0.0271 | 0.000226 | -- | 0.01 | N6009M | 06/21/2002 JDC |
| **Client ID:** ATC-0614-M894-Hg4 | **Lab ID:** -04A | **Date Sampled:** 6/14/2002 | | **Matrix:** HTBE | | **Air Vol.(L):** 120 |
| Mercury | 0.0256 | 0.000213 | -- | 0.01 | N6009M | 06/21/2002 JDC |
| **Client ID:** ATC-0614-M894-Hg5 | **Lab ID:** -05A | **Date Sampled:** 6/14/2002 | | **Matrix:** HTBE | | **Air Vol.(L):** NA |
| Mercury | 0.0248 | -- | -- | 0.01 | N6009M | 06/21/2002 JDC |

(a) Analysis indicates possible breakthrough; back section result is greater than 0% of the front section result.

General Notes:
<: Less than the indicated limit of detection (LOD).
--: Information not available or not applicable.
Back sections were checked and showed no significant breakthrough.



CITY_21MC100_002895150

NYC-WTC_000226236