NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	Volume (liters):	1810
Lab Group #:	6375	Filter Type / Filter Area:	MCE 385 mm²
Lab Sample #:	6	Grid Opening Size:	0.01124
Client Sample #:	06	Area Examined:	0.04496 mm²
GRID BOX:	75 F	Magnification:	20,000
GRID ID:	1626	Accelerating Voltage:	100 KeV
# Grid Openings:	4	Analysis Date:	1/19/02

Analysis Performed By: _____

MARK PEYSAKHOV

GO		Structure No	Structure Type				Length		ID of Structure N or √				EDS (Nonadb Elements)
No	Coord		F	B	C	M	>=05 µ <5µ	>=5µ	Chrys	Amphibole	Nonasb	Ambig/ND	
1	F7	NSD											
16 2	K7	—											
3	F3	—											
26 4	M3	—											

Struct = > 0.5µ > 5:1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart

C = Cluster (>3 Intersections)

M = Matrix

NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES

Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

NYC-WTC_000226324