

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	Volume (liters):	1936
Lab Group #:	7275	Filter Type / Filter Area:	MCE 385 mm ²
Lab Sample #:	4	Grid Opening Size:	0.01124
Client Sample #:	06	Area Examined:	0.04496 mm ²
GRID BOX:	991	Magnification:	20,000
GRID ID:	576K	Accelerating Voltage:	100 KeV
# Grid Openings:	4	Analysis Date:	3/31/02

Analysis Performed By: RP
ROMAN PEYSAKHOV

5T
6K

GO.		Structure No	Structure Type				Length		ID of Structure N or ✓				EDS (Nonadb Elements)
No	Coord		F	B	C	M	>=05 μ <5μ	>=5μ	Chrys	Amphibole	Nonasb	Ambig/ND	
1	2	NSP											
2	2	NSP											
3	2	NSP											
4	2	NSP											
5	2	NSP											
6	2	NSP											
7	2	NSP											
8	2	NSP											
9	2	NSP											
10	2	NSP											
11	2	NSP											
12	2	NSP											
13	2	NSP											
14	2	NSP											
15	2	NSP											
16	2	NSP											
17	2	NSP											
18	2	NSP											
19	2	NSP											
20	2	NSP											
21	2	NSP											
22	2	NSP											
23	2	NSP											
24	2	NSP											
25	2	NSP											
26	2	NSP											
27	2	NSP											
28	2	NSP											
29	2	NSP											
30	2	NSP											
31	2	NSP											
32	2	NSP											
33	2	NSP											
34	2	NSP											
35	2	NSP											
36	2	NSP											
37	2	NSP											
38	2	NSP											
39	2	NSP											
40	2	NSP											
41	2	NSP											
42	2	NSP											
43	2	NSP											
44	2	NSP											
45	2	NSP											
46	2	NSP											
47	2	NSP											
48	2	NSP											
49	2	NSP											
50	2	NSP											

Struct = > 0.5μ > 5:1 Aspect Ratio

✓ = Positive diffraction pattern properties conforming to laboratory identification criteria observed

✓ = EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart)

C = Cluster (>3 Intersections)

M = Matrix

NSD = No Structures Detected

Grid acceptance:

<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage at least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES

Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5μ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					