



ANALYTICAL SERVICES

INDUSTRIAL HYGIENE

ENVIRONMENTAL TESTING

• EPA NVLAP 101262-0

• AIHA ACCREDITATION NO. 100439

• NC DENR 699

• NY DOHMH LAC 10903

• PA DEP 06-353

• NJ DEP PA020

• CT DPH PH-0238

Analytical Results
for

ATC Associates

WorkOrder: R02030376

Client Reference: 19125-0297

Analyte	Concentration			Limit of Detection (ug)	Test Method	Date Analyzed /Analyst
	(ug)	(mg/m³)	(ppm)			
Client ID: HG1 ATC-0307-89M	Lab ID: -01A	Date Sampled: 3/7/02	Matrix: HTBE	Air Vol.(L): 120		
Mercury <i>1st floor hallway</i>	0.0370	0.000308	--	0.01	N6009M	03/14/2002 JDC
Client ID: HG2 ATC-0307-89M	Lab ID: -02A	Date Sampled: 3/8/02	Matrix: HTBE	Air Vol.(L): 120		
Mercury <i>Playground - north</i>	0.0360	0.000300	--	0.01	N6009M	03/14/2002 JDC
Client ID: HG3 ATC-0307-89M	Lab ID: -03A	Date Sampled: 3/8/02	Matrix: HTBE	Air Vol.(L): 120		
Mercury <i>Playground - south</i>	0.0370	0.000308	--	0.01	N6009M	03/14/2002 JDC
Client ID: HG4 ATC-0307-89M	Lab ID: -04A	Date Sampled: 3/8/02	Matrix: HTBE	Air Vol.(L): 120		
Mercury <i>Roof</i>	0.0370	0.000308	--	0.01	N6009M	03/14/2002 JDC
Client ID: HG5 ATC-0307-89M	Lab ID: -05A	Date Sampled: 3/8/02	Matrix: HTBE	Air Vol.(L): NA		
Mercury <i>Blank</i>	0.0360	--	--	0.01	N6009M	03/14/2002 JDC

(a) Analysis indicates possible breakthrough; back section result is greater than 0% of the front section result.

General Notes:

<: Less than the indicated limit of detection (LOD).

--: Information not available or not applicable.

Back sections were checked and showed no significant breakthrough.