TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	**Volume (liters):**	1944
Lab Group #:	7412	**Filter Type : Filter Area:**	MCE 385 mm²
Lab Sample #:	8	**Grid Opening Size:**	0.01124
Client Sample #:	08	**Area Examined:**	0.04496 mm²
GRID BOX:	1036	**Magnification:**	20,000
GRID ID:	1E2-c	**Accelerating Voltage:**	100 KeV
# Grid Openings:	4	**Analysis Date:**	4 14 02

Analysis Performed By: [signature]
MARK PEYSAKHOV

GO No	GO Coord	Structure No	F	B	C	M	Length >=0.5 μ <5μ	Length >=5μ	Chrys	Amphibole	Nonasb	Ambig/ND	EDS (Nonasb Elements)
1	N7	NSD											
2	E7	—											
3	D1	—											
4	U1	—											

1E
2-c

Struct = > 0.5μ > 5:1 Aspect Ratio

√ = Positive diffraction pattern properties conforming to laboratory identification criteria observed

√ = EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber
B = Bundle (<3 fibers <1 fiber diameter apart
C = Cluster (>3 Intersections)
M = Matrix
NSD = No Structures Detected

Grid acceptance:
<50% undissolved filter, overlapped or folded
<10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES
Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5u
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

CITY_21MC100_002900914