

AMBIENT GROUP, INC.

10 Morris Avenue, Glen Cove, NY 11542
 Phone: (516) 671-2370 - Fax: (516) 609-8271
 NY ELAP ID#10865

Airborne Asbestos Fiber Analysis by Transmission Electron Microscopy

Client: Lower Manhattan Development Corp.
 Address: One Liberty Plaza, 20th Floor
 New York, NY 10006
 Collected By: John Boadi
 Collected: 12/12/2004 6:02
 Rotometer: Standard


Job Number: 2043
 Date Received: 12/13/2004
 Date Analyzed: 12/16/2004
 Job Location: 130 Liberty Street

Lab ID Field ID	Sample Location	Start/ Stop	Flow	Volume (liters)	Asbestos Type(s)	Asbestos Structures		Non- Asbestos	Analytical Sensitivity (AS/cc)	Asbestos Concentration (AS/cc)	Asbestos Concentration (AS/mm2)
						≥ 5µm <5µm	≥ 5µm				
0013423-001 27607-7	Washington Street/ Albany Street	6:02 14:06	4	1936	None Detected	0	0	0	0.0008	<0.0008	<3.98
0013423-002 27607-8	Albany Street/ Greenwich Street	6:04 14:09	4	1940	None Detected	0	0	0	0.0008	<0.0008	<3.98
0013423-003 27607-9	Greenwich Street	6:09 14:14	4	1940	None Detected	0	0	0	0.0008	<0.0008	<3.98
0013423-004 27607-10	Liberty Street	6:16 14:20	4	1936	None Detected	0	0	0	0.0008	<0.0008	<3.98
0013423-005 27607-11	Washington Street/ Cedar Street	6:22 14:24	3.75	1807.5	None Detected	0	0	0	0.0008	<0.0008	<3.79

SIGNED FOR AMBIENT GROUP, INC. BY:

Page 1 of 3

Disclaimers: The laboratory is not responsible for fibers counted in fibers/cc which is dependent on volume collected by non-laboratory personnel.
 This report may not be duplicated in part without written permission by Ambient Group, Inc. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government.
 This report relates only to the samples listed above.



AMBIENT GROUP, INC.

10 Morris Avenue, Glen Cove, NY 11542
 Phone: (516) 671-2370 - Fax: (516) 609-8271
 NY ELAP ID#10865

Airborne Asbestos Fiber Analysis by Transmission Electron Microscopy

Client: Lower Manhattan Development Corp.
 Address: One Liberty Plaza, 20th Floor
 New York, NY 10006
 Collected By: William Kennedy
 Collected: 12/12/2004 14:06
 Rotometer: Standard


Job Number: 2043
 Date Received: 12/13/2004
 Date Analyzed: 12/16/2004
 Job Location: 130 Liberty Street

Lab ID						Asbestos Structures		Non-Asbestos	Analytical Sensitivity (AS/cc)	Asbestos Concentration (AS/cc)	Asbestos Concentration (AS/mm2)
Field ID	Sample Location	Start/Stop	Flow	Volume (liters)	Asbestos Type(s)	≥ 5µm <5µm	≥ 5µm				
0013424-001	Washington Street/ Albany Street	14:06 22:02	3.5	1666	None Detected	0	0	0	0.0008	<0.0008	<3.32
27616-7											
0013424-002	Albany Street/ Greenwich Street	14:09 22:05	3.75	1785	None Detected	0	0	0	0.0008	<0.0008	<3.62
27616-8											
0013424-003	Greenwich Street	14:14 22:10	3.5	1666	None Detected	0	0	0	0.0008	<0.0008	<3.32
27616-9											
0013424-004	Liberty Street	14:20 22:14	4.5	2133	None Detected	0	0	0	0.0008	<0.0008	<4.43
27616-10											
0013424-005	Washington Street/ Cedar Street	14:24 22:20	4	1904	None Detected	0	0	0	0.0008	<0.0008	<3.98
27616-11											

SIGNED FOR AMBIENT GROUP, INC. BY:

Page 2 of 3

Disclaimers: The laboratory is not responsible for fibers counted in fibers/cc which is dependent on volume collected by non-laboratory personnel.
 This report may not be duplicated in part without written permission by Ambient Group, Inc. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government.
 This report relates only to the samples listed above.



AMBIENT GROUP, INC.

10 Morris Avenue, Glen Cove, NY 11542
 Phone: (516) 671-2370 - Fax: (516) 609-8271
 NY ELAP ID#10865

Airborne Asbestos Fiber Analysis by Transmission Electron Microscopy

Client: Lower Manhattan Development Corp.
 Address: One Liberty Plaza, 20th Floor
 New York, NY 10006
 Collected By: Rotimi Adeola
 Collected: 12/12/2004 22:02
 Rotometer: Standard

Job Number: 2043
 Date Received: 12/13/2004
 Date Analyzed: 12/16/2004
 Job Location: 130 Liberty Street

Lab ID						Asbestos Structures		Non-Asbestos	Analytical Sensitivity (AS/cc)	Asbestos Concentration (AS/cc)	Asbestos Concentration (AS/mm2)
Field ID	Sample Location	Start/Stop	Flow	Volume (liters)	Asbestos Type(s)	≥ 5µm <5µm	≥ 5µm				
0013425-001	Washington Street/ Albany Street	22:02 6:13	4.5	2209.5	None Detected	0	0	0	0.0008	<0.0008	<4.69
27626-7											
0013425-002	Albany Street/ Greenwich Street	22:05 6:20	4	1980	None Detected	0	0	0	0.0008	<0.0008	<3.98
27626-8											
0013425-003	Greenwich Street	22:10 6:25	4	1980	None Detected	0	0	0	0.0008	<0.0008	<3.98
27626-9											
0013425-004	Liberty Street	22:14 6:37	4	2012	None Detected	0	0	0	0.0008	<0.0008	<4.19
27626-10											
0013425-005	Washington Street/ Cedar Street	22:20 6:41	4.5	2254.5	None Detected	0	0	0	0.0008	<0.0008	<4.69
27626-11											

Average Filter Loading: <3.98

SIGNED FOR AMBIENT GROUP, INC. BY:

Page 3 of 3

Disclaimers: The laboratory is not responsible for fibers counted in fibers/cc which is dependent on volume collected by non-laboratory personnel.
 This report may not be duplicated in part without written permission by Ambient Group, Inc. This report must not be used to claim product endorsement by NVLAP or any agency of the U.S. Government.
 This report relates only to the samples listed above.