

**Galson
Laboratories**

6601 Kirkville Rd. E. Syracuse, NY 13057

LABORATORY ANALYSIS REPORT

Client : HP Environmental, Inc.
Site : NS
Project No. : HPE-01

Date Sampled : 02-OCT-01
Date Received : 05-OCT-01
Date Analyzed : 05-OCT-01

Account No.: 14646
Login No. : L75338

Silica: Quartz, Cristobalite & Tridymite

<u>Sample ID</u>	<u>Lab ID</u>	<u>Air Vol</u> <u>m3</u>	<u>Quartz</u> <u>mg</u>	<u>Quartz</u> <u>mg/m3</u>	<u>Cristo</u> <u>mg</u>	<u>Cristo</u> <u>mg/m3</u>	<u>Tridym</u> <u>mg</u>	<u>Tridym</u> <u>mg/m3</u>
WFC1002-ASB-A2	L75338-1	0.566	<0.01	<0.02	NA	NA	NA	NA
WFC1002-ASB-P1	L75338-2	0.638	<0.01	<0.02	NA	NA	NA	NA
LAB BLANK	L75338-3	NA	<0.01	NA	NA	NA	NA	NA

COMMENTS: Samples collected on MCE filters, NIOSH method 7500 requires PVC filters.

Level of quantitation: Silica 0.01mg
Analytical Method : NIOSH 7500
OSHA PEL (TWA) : NA
Collection Media : Filter

Submitted by: AP
Approved by : DM
Date : 05-OCT-01
QC by: QC STAFF
NYS DOH # : 11626

< -Less Than	mg -Milligrams	m3 -Cubic Meters	kg -Kilograms
> -Greater Than	ug -Micrograms	l -Liters	NS -Not Specified
NA -Not Applicable	ND -Not Detected	ppm -Parts per Million	

page 1 of 1

06.002496

CITY_21MC100_003077413