NYC 9/11 Public Portal Document

ASBESTOS AIR SAMPLING CALIBRATION SCHEDULE

INTERNATIONAL TECHNOLOGY CORPORATION

DATE: 10/01/01

Pump ID	Sample ID Number	Location	Pre-Cal. (lpm)	Post-Cal. (lpm)	Time (on/off)	Total Volume	Results (f/cc)
#SKC-05	WTC-100104-A1	TURNER TRAILER ENTRANCE WAY	2.0	2.0	1700/1105	490	0.039
#842	WTC-100101-P1	TURNER employee - CARL	2.0	2.0	0723/1300	674	<0.008
#SKC-06	WTC-100101-A2	ET TURNER TRAILER NEXT TO BLDG 7 INTERIOR	2.0	2.0	0730/1420	820	<0.005

Calsible.wpd/11-97

SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008 CONFIDENTIAL,

06.002595

CITY_21MC100_003416596

NYC-WTC_000237828