NYC 9/11 Public Portal Document

PSC ANALYTICAL SERVICES

INDUSTRIAL HYGIENE ENVIRONMENTAL TESTING

• EPA/NVLAP 101262-0 • NY DOH/NELAC 10903 • NJ DEP PA020
• AIHA ACCREDITATION NO. 100439 • PA DEP 06-353 • CT DPH PH-0238
• NC DENR 599

Analytical Results
for
ATC Associates

WorkOrder: R02090386
Client Reference: Murry Bergtraum HS

Analyte	Concentration (ug)	Concentration (mg/m³)	Concentration (ppm)	Limit of Detection (ug)	Test Method	Date Analyzed /Analyst	
Client ID: ATC-0911-M520-M1	**Lab ID:** -01A	**Date Sampled:** 9/11/2002		**Matrix:** Filter		**Air Vol.(L):** 1200	
Aluminum	<1.00	<0.000833	--	1	N7300 MCE	09/14/2002	AJB
Cadmium	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Chromium	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Cobalt	<0.100	<0.0000833	--	0.1	N7300 MCE	09/14/2002	AJB
Copper	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Iron	<2.00	<0.00167	--	2	N7300 MCE	09/14/2002	AJB
Lead	<0.300	<0.000250	--	0.3	N7300 MCE	09/14/2002	AJB
Magnesium	<1.30	<0.00108	--	1.3	N7300 MCE	09/14/2002	AJB
Manganese	<0.300	<0.000250	--	0.3	N7300 MCE	09/14/2002	AJB
Molybdenum	<1.00	<0.000833	--	1	N7300 MCE	09/14/2002	AJB
Nickel	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Zinc	<0.500	<0.000417	--	0.5	N7300 MCE	09/14/2002	AJB
Client ID: ATC-0911-M520-M2	**Lab ID:** -02A	**Date Sampled:** 9/11/2002		**Matrix:** Filter		**Air Vol.(L):** 1200	
Aluminum	1.02	0.000853	--	1	N7300 MCE	09/14/2002	AJB
Cadmium	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Chromium	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Cobalt	<0.100	<0.0000833	--	0.1	N7300 MCE	09/14/2002	AJB
Copper	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Iron	4.42	0.00368	--	2	N7300 MCE	09/14/2002	AJB
Lead	<0.300	<0.000250	--	0.3	N7300 MCE	09/14/2002	AJB
Magnesium	<1.30	<0.00108	--	1.3	N7300 MCE	09/14/2002	AJB
Manganese	<0.300	<0.000250	--	0.3	N7300 MCE	09/14/2002	AJB
Molybdenum	<1.00	<0.000833	--	1	N7300 MCE	09/14/2002	AJB
Nickel	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Zinc	<0.500	<0.000417	--	0.5	N7300 MCE	09/14/2002	AJB

CONFIDENTIAL, CITY_21MC100_003422332

SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008

NYC-WTC_000239355

NYC 9/11 Public Portal Document

PSC ANALYTICAL SERVICES

INDUSTRIAL HYGIENE ENVIRONMENTAL TESTING

- EPA/NVLAP 101262-0
- AIHA ACCREDITATION NO. 100439
- NC DENR 599
- NY DOH/NELAC 10903
- PA DEP 06-353
- NJ DEP PA020
- CT DPH PH-0238

Analytical Results
for
ATC Associates

WorkOrder: R02090386
Client Reference: Murry Bergtraum HS

Analyte	Concentration (ug)	Concentration (mg/m³)	Concentration (ppm)	Limit of Detection (ug)	Test Method	Date Analyzed /Analyst	
Client ID: ATC-0911-M520-M3	**Lab ID:** -03A	**Date Sampled:** 9/11/2002		**Matrix:** Filter		**Air Vol.(L):** 1200	
Aluminum	<1.00	<0.000833	--	1	N7300 MCE	09/14/2002	AJB
Cadmium	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Chromium	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Cobalt	<0.100	<0.0000833	--	0.1	N7300 MCE	09/14/2002	AJB
Copper	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Iron	<2.00	<0.00167	--	2	N7300 MCE	09/14/2002	AJB
Lead	<0.300	<0.000250	--	0.3	N7300 MCE	09/14/2002	AJB
Magnesium	<1.30	<0.00108	--	1.3	N7300 MCE	09/14/2002	AJB
Manganese	<0.300	<0.000250	--	0.3	N7300 MCE	09/14/2002	AJB
Molybdenum	<1.00	<0.000833	--	1	N7300 MCE	09/14/2002	AJB
Nickel	<0.200	<0.000167	--	0.2	N7300 MCE	09/14/2002	AJB
Zinc	<0.500	<0.000417	--	0.5	N7300 MCE	09/14/2002	AJB
Client ID: ATC-0911-M520-M4	**Lab ID:** -04A	**Date Sampled:** 9/11/2002		**Matrix:** Filter		**Air Vol.(L):** NA	
Aluminum	<1.00	--	--	1	N7300 MCE	09/14/2002	AJB
Cadmium	<0.200	--	--	0.2	N7300 MCE	09/14/2002	AJB
Chromium	0.460	--	--	0.2	N7300 MCE	09/14/2002	AJB
Cobalt	<0.100	--	--	0.1	N7300 MCE	09/14/2002	AJB
Copper	<0.200	--	--	0.2	N7300 MCE	09/14/2002	AJB
Iron	<2.00	--	--	2	N7300 MCE	09/14/2002	AJB
Lead	<0.300	--	--	0.3	N7300 MCE	09/14/2002	AJB
Magnesium	<1.30	--	--	1.3	N7300 MCE	09/14/2002	AJB
Manganese	<0.300	--	--	0.3	N7300 MCE	09/14/2002	AJB
Molybdenum	<1.00	--	--	1	N7300 MCE	09/14/2002	AJB
Nickel	<0.200	--	--	0.2	N7300 MCE	09/14/2002	AJB
Zinc	<0.500	--	--	0.5	N7300 MCE	09/14/2002	AJB

CONFIDENTIAL, CITY_21MC100_003422333

SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008

NYC-WTC_000239356

NYC 9/11 Public Portal Document

PSC ANALYTICAL SERVICES

INDUSTRIAL HYGIENE ENVIRONMENTAL TESTING

• EPA/NVLAP 101262-0 • NY DOH/NELAC 10903 • NJ DEP PA020
• AIHA ACCREDITATION NO. 100439 • PA DEP 06-353 • CT DPH PH-0238
• NC DENR 599

Analytical Results
for
ATC Associates

WorkOrder: R02090386
Client Reference: Murry Bergtraum HS

Analyte	Concentration (ug)	Concentration (mg/m³)	Concentration (ppm)	Limit of Detection (ug)	Test Method	Date Analyzed /Analyst	
Client ID: ATC-0911-M520-M5	**Lab ID:** -05A	**Date Sampled:** 9/11/2002		**Matrix:** Filter		**Air Vol.(L):** NA	
Aluminum	<1.00	--	--	1	N7300 MCE	09/14/2002	AJB
Cadmium	<0.200	--	--	0.2	N7300 MCE	09/14/2002	AJB
Chromium	<0.200	--	--	0.2	N7300 MCE	09/14/2002	AJB
Cobalt	<0.100	--	--	0.1	N7300 MCE	09/14/2002	AJB
Copper	<0.200	--	--	0.2	N7300 MCE	09/14/2002	AJB
Iron	<2.00	--	--	2	N7300 MCE	09/14/2002	AJB
Lead	<0.300	--	--	0.3	N7300 MCE	09/14/2002	AJB
Magnesium	<1.30	--	--	1.3	N7300 MCE	09/14/2002	AJB
Manganese	<0.300	--	--	0.3	N7300 MCE	09/14/2002	AJB
Molybdenum	<1.00	--	--	1	N7300 MCE	09/14/2002	AJB
Nickel	<0.200	--	--	0.2	N7300 MCE	09/14/2002	AJB
Zinc	<0.500	--	--	0.5	N7300 MCE	09/14/2002	AJB

(a) Analysis indicates possible breakthrough; back section result is greater than 0% of the front section result.

General Notes:
<: Less than the indicated limit of detection (LOD).
--: Information not available or not applicable.
Back sections were checked and showed no significant breakthrough.

CONFIDENTIAL, CITY_21MC100_003422334
SUBJECT TO PROTECTIVE ORDER NO. 2, SEPT. 2008

NYC-WTC_000239357