NYC 9/11 Public Portal Document

ATHENICA
ENVIRONMENTAL SERVICES INC.

45-09 Greenpoint Avenue
Long Island City, N.Y. 11104

Tel. (718) 784-7490
Fax (718) 784-4085

## TEM DEPARTMENT

## TEM AHERA COUNTSHEET

| | |
|---|---|
| CLIENT : Warren & Panzer | DATE : 8/1/03 |
| PROJECT : MCDEP/15 Park Row | ANALYST : ES |
| BATCH # : T03-02-001 | CLIENT ID # : 02 |
| SAMPLE ID # : T03-02-001-125 | VOLUME (lt) : 1680 |
| G.O. AREA ($mm^2$) : 0.012820 | GO's TO ANALYZE : 4 |
| MAGNIFICATION : 194 | ACC. VOLTAGE (KeV) : 100 |
| INSTRUMENT : JEOL 100CX II | FILTER AREA ($mm^2$) : 385. |
| ANALYTICAL METHODOLOGY: EPA FEDERAL REGISTER, VOL. 52, No 210,OCT. 30, 1987, #40 CFR, PART 763, APPENDIX A TO SUBPART E | |

| GRID | GO ID | Str. # | Str. Classf. | L (µm) | W (µm) | Structure sketch | SAED | EDXA | PHOTO# | ID |
|---|---|---|---|---|---|---|---|---|---|---|
| B | D8 | NSD | | | | | | | | |
| | B5 | NSD | | | | | | | | |
| A | E4 | NSD | | | | | | | | |
| | H7 | NSD | | | | | | | | |
| | | | | | | | | | | |
| | | | | | | | | | | |
| | | | | | | | | | | |
| | | | | | | | | | | |
| | | | | | | | | | | |
| | | | | | | | | | | |

| | |
|---|---|
| Analytical Sensitivity (str/$cm^3$): 0.0045 | Eucentricity check : |
| Concentration (str/$cm^3$) : <0.0045 | *Grid acceptance criteria check : (A) ✓ (B) ✓ |
| Concentration (str/$mm^2$) : <19.50 | **CHR. # STRUCT. : F M C B |
| Number of Asbestos Structures : 0.5≤S<5µm ...0. | **AMPH. # STRUCT.:F M C B |
| S≥5µm .....0. | Specify AMPH. Type : AM. ACT. TRE. |
| Number of Non-asbestos Structures:.....0. | CRO. ANTH. |

*Replica covers >50%,at least 20 GO's with no overlapping or folded replica, intact GO's have >50% coverage, <50% Undissolved filter, <5% holes or opaque areas, <10% particulate coverage.

** CHR =chrysotile, AMPH.=amphibole, AM.=amosite, ACT.=actinolite, TRE.=tremolite, CRO.=crocidolite, ANTH.=anthophyllite



NYC-WTC_000282951