NYC 9/11 Public Portal Document

TEM Asbestos (air) Count Sheet

Client Name:	NYC BOE	**Volume (liters).**	1880
Lab Group #:	6831	**Filter Type / Filter Area**	MCE 385 mm²
Lab Sample #:	15	**Grid Opening Size.**	0.01124
Client Sample #:	T18	**Area Examined:**	0.04496 mm²
GRID BOX:	E 6	**Magnification:**	20,000
GRID ID:	E 13	**Accelerating Voltage:**	100 KeV
# Grid Openings:	4	**Analysis Date:**	2/25/02

Analysis Performed By: MARK PEYSAKHOV

GO	Structure	Structure Type				Length		ID of Structure N or Y				EDS
No Coord	No	F	B	C	M	>=05 µ <5µ	>=5µ	Chrys	Amphibole	Nonasb	Ambig/ND	(Nonadb Elements)
1 E 3	1											
2 J 3	—											
3 C 7	-											
4 C 7	-											

Struct = > 0 5µ > 5 1 Aspect Ratio

√= Positive diffraction pattern properties conforming to laboratory identification criteria observed

√= EDS spectrum conforming to laboratory identification criteria observed

F = Single Fiber

B = Bundle (<3 fibers <1 fiber diameter apart

C = Cluster (>3 Intersections)

M = Matrix

NSD = No Structures Detected

Grid acceptance

<50% undissolved filter, overlapped or folded <10% particulate coverage >50% coverage ate least 20 unfolded or overlapped grid squares <5% holes or opaque areas

AMPHIBOLES

Amosite
Crocidolite
Anthophyllite
Tremolite
Actinolite

Totals

	F	B	C	M	>5µ
Chrysotile					
Amphibole					
Ambiguous					
Non-Asbestos					
Non-Identified					

NYC-WTC_000298076