NYC 9/11 Public Portal Document

PSC ANALYTICAL SERVICES

INDUSTRIAL HYGIENE ENVIRONMENTAL TESTING

- EPA/NVLAP 101262-0
- AIHA ACCREDITATION NO. 100439
- NC DENR 699
- NY DOH/NELAC 10903
- PA DEP 06-353
- NJ DEP PA020
- CT DPH PH-0238

Analytical Results
for
ATC Associates

WorkOrder: R02060659
Client Reference: 19125-0289, Metals

Analyte	Concentration (ug)	Concentration (mg/m³)	Concentration (ppm)	Limit of Detection (ug)	Test Method	Date Analyzed /Analyst	
Client ID: ATC-0617-150M-M1	**Lab ID:** -01A	**Date Sampled:** 6/17/2002		**Matrix:** Filter		**Air Vol.(L):** 1200	
Aluminum	<1.00	<0.000833	--	1	N7300 MCE	06/20/2002	AJB
Cadmium	<0.200	<0.000167	--	0.2	N7300 MCE	06/20/2002	AJB
Chromium	0.316	0.000264	--	0.2	N7300 MCE	06/20/2002	AJB
Cobalt	<0.100	<0.0000833	--	0.1	N7300 MCE	06/20/2002	AJB
Copper	<0.200	<0.000167	--	0.2	N7300 MCE	06/20/2002	AJB
Iron	<2.00	<0.00167	--	2	N7300 MCE	06/20/2002	AJB
Lead	<0.300	<0.000250	--	0.3	N7300 MCE	06/20/2002	AJB
Magnesium	<1.30	<0.00108	--	1.3	N7300 MCE	06/20/2002	AJB
Manganese	<0.300	<0.000250	--	0.3	N7300 MCE	06/20/2002	AJB
Molybdenum	<1.00	<0.000833	--	1	N7300 MCE	06/20/2002	AJB
Nickel	<0.200	<0.000167	--	0.2	N7300 MCE	06/20/2002	AJB
Zinc	<0.500	<0.000417	--	0.5	N7300 MCE	06/20/2002	AJB
Client ID: ATC-0617-150M-M2	**Lab ID:** -02A	**Date Sampled:** 6/17/2002		**Matrix:** Filter		**Air Vol.(L):** 1200	
Aluminum	1.04	0.000864	--	1	N7300 MCE	06/20/2002	AJB
Cadmium	<0.200	<0.000167	--	0.2	N7300 MCE	06/20/2002	AJB
Chromium	0.328	0.000274	--	0.2	N7300 MCE	06/20/2002	AJB
Cobalt	<0.100	<0.0000833	--	0.1	N7300 MCE	06/20/2002	AJB
Copper	<0.200	<0.000167	--	0.2	N7300 MCE	06/20/2002	AJB
Iron	<2.00	<0.00167	--	2	N7300 MCE	06/20/2002	AJB
Lead	<0.300	<0.000250	--	0.3	N7300 MCE	06/20/2002	AJB
Magnesium	<1.30	<0.00108	--	1.3	N7300 MCE	06/20/2002	AJB
Manganese	<0.300	<0.000250	--	0.3	N7300 MCE	06/20/2002	AJB
Molybdenum	<1.00	<0.000833	--	1	N7300 MCE	06/20/2002	AJB
Nickel	<0.200	<0.000167	--	0.2	N7300 MCE	06/20/2002	AJB
Zinc	<0.500	<0.000417	--	0.5	N7300 MCE	06/20/2002	AJB

CITY_21MC100_004152418

NYC-WTC_000301034

NYC 9/11 Public Portal Document

PSC ANALYTICAL SERVICES

INDUSTRIAL HYGIENE ENVIRONMENTAL TESTING

- EPA/NVLAP 101262-0
- AIHA ACCREDITATION NO. 100439
- NC DENR 599
- NY DOH/NELAC 10903
- PA DEP 06-353
- NJ DEP PA020
- CT DPH PH-0238

Analytical Results
for
ATC Associates

WorkOrder: R02060659
Client Reference: 19125-0289, Metals

Analyte	Concentration (ug)	Concentration (mg/m³)	Concentration (ppm)	Limit of Detection (ug)	Test Method	Date Analyzed /Analyst	
Client ID: ATC-0617-150M-M3	**Lab ID:** -03A	**Date Sampled:** 6/17/2002		**Matrix:** Filter		**Air Vol.(L):** 1200	
Aluminum	<1.00	<0.000833	--	1	N7300 MCE	06/20/2002	AJB
Cadmium	<0.200	<0.000167	--	0.2	N7300 MCE	06/20/2002	AJB
Chromium	0.337	0.000281	--	0.2	N7300 MCE	06/20/2002	AJB
Cobalt	<0.100	<0.0000833	--	0.1	N7300 MCE	06/20/2002	AJB
Copper	<0.200	<0.000167	--	0.2	N7300 MCE	06/20/2002	AJB
Iron	<2.00	<0.00167	--	2	N7300 MCE	06/20/2002	AJB
Lead	<0.300	<0.000250	--	0.3	N7300 MCE	06/20/2002	AJB
Magnesium	<1.30	<0.00108	--	1.3	N7300 MCE	06/20/2002	AJB
Manganese	<0.300	<0.000250	--	0.3	N7300 MCE	06/20/2002	AJB
Molybdenum	<1.00	<0.000833	--	1	N7300 MCE	06/20/2002	AJB
Nickel	<0.200	<0.000167	--	0.2	N7300 MCE	06/20/2002	AJB
Zinc	<0.500	<0.000417	--	0.5	N7300 MCE	06/20/2002	AJB
Client ID: ATC-0617-150M-M4	**Lab ID:** -04A	**Date Sampled:** 6/17/2002		**Matrix:** Filter		**Air Vol.(L):** NA	
Aluminum	<1.00	--	--	1	N7300 MCE	06/20/2002	AJB
Cadmium	<0.200	--	--	0.2	N7300 MCE	06/20/2002	AJB
Chromium	0.335	--	--	0.2	N7300 MCE	06/20/2002	AJB
Cobalt	<0.100	--	--	0.1	N7300 MCE	06/20/2002	AJB
Copper	<0.200	--	--	0.2	N7300 MCE	06/20/2002	AJB
Iron	<2.00	--	--	2	N7300 MCE	06/20/2002	AJB
Lead	<0.300	--	--	0.3	N7300 MCE	06/20/2002	AJB
Magnesium	<1.30	--	--	1.3	N7300 MCE	06/20/2002	AJB
Manganese	<0.300	--	--	0.3	N7300 MCE	06/20/2002	AJB
Molybdenum	<1.00	--	--	1	N7300 MCE	06/20/2002	AJB
Nickel	<0.200	--	--	0.2	N7300 MCE	06/20/2002	AJB
Zinc	<0.500	--	--	0.5	N7300 MCE	06/20/2002	AJB

CITY_21MC100_004152419

NYC-WTC_000301035

NYC 9/11 Public Portal Document

PSC ANALYTICAL SERVICES

INDUSTRIAL HYGIENE ENVIRONMENTAL TESTING

- EPA/NVLAP 101262-0
- AIHA ACCREDITATION NO. 100439
- NC DENR 599
- NY DOH/NELAC 10903
- PA DEP 06-353
- NJ DEP PA020
- CT DPH PH-0238

Analytical Results
for
ATC Associates

WorkOrder: R02060659
Client Reference: 19125-0289, Metals

Analyte	Concentration (ug)	Concentration (mg/m³)	Concentration (ppm)	Limit of Detection (ug)	Test Method	Date Analyzed /Analyst	
Client ID: ATC-0617-150M-M5	**Lab ID:** -05A	**Date Sampled:** 6/17/2002		**Matrix:** Filter		**Air Vol.(L):** NA	
Aluminum	1.01	--	--	1	N7300 MCE	06/20/2002	AJB
Cadmium	<0.200	--	--	0.2	N7300 MCE	06/20/2002	AJB
Chromium	0.310	--	--	0.2	N7300 MCE	06/20/2002	AJB
Cobalt	<0.100	--	--	0.1	N7300 MCE	06/20/2002	AJB
Copper	<0.200	--	--	0.2	N7300 MCE	06/20/2002	AJB
Iron	<2.00	--	--	2	N7300 MCE	06/20/2002	AJB
Lead	<0.300	--	--	0.3	N7300 MCE	06/20/2002	AJB
Magnesium	<1.30	--	--	1.3	N7300 MCE	06/20/2002	AJB
Manganese	<0.300	--	--	0.3	N7300 MCE	06/20/2002	AJB
Molybdenum	<1.00	--	--	1	N7300 MCE	06/20/2002	AJB
Nickel	<0.200	--	--	0.2	N7300 MCE	06/20/2002	AJB
Zinc	<0.500	--	--	0.5	N7300 MCE	06/20/2002	AJB

(a) Analysis indicates possible breakthrough; back section result is greater than 0% of the front section result.

General Notes:
< Less than the indicated limit of detection (LOD).
-- Information not available or not applicable.
Back sections were checked and showed no significant breakthrough.

CITY_21MC100_004152420

NYC-WTC_000301036